

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX082918\
 Data File : VX004310.D
 Acq On : 29 Aug 2018 23:51
 Operator : JC/MD
 Sample : J4681-18
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPA-SW-15-M12W4-082218

Quant Time: Aug 30 07:31:21 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	262301	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	410167	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	390365	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208569	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	186387	54.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.04%	
35) Dibromofluoromethane	5.50	113	163985	53.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.48%	
50) Toluene-d8	8.72	98	593446	51.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.04%	
62) 4-Bromofluorobenzene	11.14	95	205538	47.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.66%	
Target Compounds						
17) Carbon Disulfide	2.57	76	3311	1.22	ug/l	96
27) cis-1,2-Dichloroethene	4.59	96	1182	0.34	ug/l	97
37) Ethyl Acetate	4.86	43	21058	4.64	ug/l	99
64) Tetrachloroethene	9.34	164	4132	1.07	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX082918\
Data File : VX004310.D
Acq On : 29 Aug 2018 23:51
Operator : JC/MD
Sample : J4681-18
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPA-SW-15-M12W4-082218

Quant Time: Aug 30 07:31:21 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

