


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX082919\
Data File : VX011883.D
Acq On : 29 Aug 2019 19:53
Operator : SY/SP
Sample : BFB
Misc : 5.00µ/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Title : SW846 8260
Last Update : Fri Aug 30 01:59:05 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1629

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.7	21784	PASS
75	95	30	60	48.9	60224	PASS
95	95	100	100	100.0	123197	PASS
96	95	5	9	6.6	8080	PASS
173	174	0.00	2	0.3	374	PASS
174	95	50	100	97.3	119861	PASS
175	174	5	9	7.3	8807	PASS
176	174	95	101	96.7	115933	PASS
177	176	5	9	6.8	7895	PASS