

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037360.D
 Acq On : 30 Aug 2023 01:01
 Operator : JC/MD
 Sample : VX0829WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 30 02:12:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	94152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	191512	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	185212	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91307	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88503	65.202	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	130.400%#
35) Dibromofluoromethane	5.385	113	73051	57.596	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.200%
50) Toluene-d8	8.647	98	243935	53.021	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.040%
62) 4-Bromofluorobenzene	11.079	95	101903	56.361	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.720%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	18059	16.512	ug/l	95
3) Chloromethane	1.300	50	23577	24.641	ug/l	98
4) Vinyl Chloride	1.374	62	21281	20.433	ug/l	97
5) Bromomethane	1.593	94	9862	19.010	ug/l	99
6) Chloroethane	1.672	64	13756	22.445	ug/l	95
7) Trichlorofluoromethane	1.880	101	29500	17.291	ug/l	100
8) Diethyl Ether	2.136	74	16428	24.889	ug/l	77
9) 1,1,2-Trichlorotrifluo...	2.325	101	15323	15.396	ug/l	93
10) Methyl Iodide	2.447	142	21533	17.801	ug/l	93
11) Tert butyl alcohol	2.977	59	41156	129.256	ug/l #	95
12) 1,1-Dichloroethene	2.312	96	17886	18.454	ug/l	85
13) Acrolein	2.239	56	17172	70.892	ug/l	98
14) Allyl chloride	2.660	41	28264	18.841	ug/l #	90
15) Acrylonitrile	3.068	53	74596	131.258	ug/l	98
16) Acetone	2.386	43	63586	119.744	ug/l #	88
17) Carbon Disulfide	2.508	76	42225	17.128	ug/l	100
18) Methyl Acetate	2.703	43	49097	24.556	ug/l #	87
19) Methyl tert-butyl Ether	3.117	73	86266	23.469	ug/l	99
20) Methylene Chloride	2.788	84	28469	23.974	ug/l	86
21) trans-1,2-Dichloroethene	3.087	96	22232	20.179	ug/l	86
22) Diisopropyl ether	3.763	45	71464	23.215	ug/l #	87
23) Vinyl Acetate	3.721	43	272244	115.407	ug/l #	92
24) 1,1-Dichloroethane	3.611	63	42711	22.125	ug/l	96
25) 2-Butanone	4.562	43	98921	123.721	ug/l #	89
26) 2,2-Dichloropropane	4.471	77	20381	11.822	ug/l	92
27) cis-1,2-Dichloroethene	4.489	96	29206	22.486	ug/l	83
28) Bromochloromethane	4.891	49	22028	25.544	ug/l #	72
29) Tetrahydrofuran	5.007	42	62598	122.766	ug/l #	83
30) Chloroform	5.092	83	48395	22.333	ug/l	91
31) Cyclohexane	5.470	56	25087	16.437	ug/l	88
32) 1,1,1-Trichloroethane	5.385	97	37107	19.109	ug/l	96
36) 1,1-Dichloropropene	5.696	75	27821	15.439	ug/l	98
37) Ethyl Acetate	4.714	43	37058	20.132	ug/l #	94
38) Carbon Tetrachloride	5.678	117	29700	15.173	ug/l	96
39) Methylcyclohexane	7.379	83	26294	13.888	ug/l	93
40) Benzene	6.037	78	98716	19.101	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037360.D
 Acq On : 30 Aug 2023 01:01
 Operator : JC/MD
 Sample : VX0829WBSD02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Aug 30 02:12:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.922	41	19290	20.159	ug/l #	84
42) 1,2-Dichloroethane	6.086	62	39779	19.628	ug/l	94
43) Isopropyl Acetate	6.342	43	57240	19.587	ug/l #	90
44) Trichloroethene	7.129	130	24089	16.438	ug/l	96
45) 1,2-Dichloropropane	7.433	63	25639	20.120	ug/l	96
46) Dibromomethane	7.580	93	20567	20.488	ug/l	95
47) Bromodichloromethane	7.824	83	36146	18.799	ug/l	100
48) Methyl methacrylate	7.696	41	28152	19.753	ug/l #	85
49) 1,4-Dioxane	7.665	88	17490	455.677	ug/l #	90
51) 4-Methyl-2-Pentanone	8.573	43	191885	108.132	ug/l	90
52) Toluene	8.720	92	64512	18.893	ug/l	98
53) t-1,3-Dichloropropene	8.982	75	34754	16.976	ug/l	98
54) cis-1,3-Dichloropropene	8.366	75	39012	17.633	ug/l #	84
55) 1,1,2-Trichloroethane	9.153	97	30220	21.734	ug/l	97
56) Ethyl methacrylate	9.116	69	44211	20.300	ug/l	86
57) 1,3-Dichloropropane	9.311	76	49488	21.113	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.244	63	120473	115.268	ug/l	92
59) 2-Hexanone	9.433	43	148606	109.831	ug/l	88
60) Dibromochloromethane	9.518	129	29101	19.658	ug/l	99
61) 1,2-Dibromoethane	9.610	107	31786	20.880	ug/l	100
64) Tetrachloroethene	9.275	164	19087	14.127	ug/l	95
65) Chlorobenzene	10.079	112	72253	18.458	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.165	131	26971	18.668	ug/l	98
67) Ethyl Benzene	10.195	91	116464	16.908	ug/l	99
68) m/p-Xylenes	10.299	106	91012	34.215	ug/l	99
69) o-Xylene	10.640	106	48091	18.325	ug/l	99
70) Styrene	10.658	104	79770	18.452	ug/l	97
71) Bromoform	10.799	173	20658	19.115	ug/l #	99
73) Isopropylbenzene	10.963	105	111970	16.871	ug/l	98
74) N-amyl acetate	10.841	43	50346	19.523	ug/l #	88
75) 1,1,2,2-Tetrachloroethane	11.213	83	48203	21.567	ug/l	99
76) 1,2,3-Trichloropropane	11.238	75	48882	25.138	ug/l	91
77) Bromobenzene	11.201	156	31071	19.170	ug/l	93
78) n-propylbenzene	11.305	91	121463	16.436	ug/l	99
79) 2-Chlorotoluene	11.366	91	85808	18.170	ug/l	98
80) 1,3,5-Trimethylbenzene	11.451	105	97288	17.573	ug/l	98
81) trans-1,4-Dichloro-2-b...	11.018	75	10133	14.555	ug/l	94
82) 4-Chlorotoluene	11.457	91	95452	17.510	ug/l	98
83) tert-Butylbenzene	11.713	119	97311	18.431	ug/l	95
84) 1,2,4-Trimethylbenzene	11.750	105	101831	18.200	ug/l	98
85) sec-Butylbenzene	11.890	105	106000	17.188	ug/l	99
86) p-Isopropyltoluene	12.012	119	91417	17.220	ug/l	98
87) 1,3-Dichlorobenzene	11.969	146	55471	18.280	ug/l	99
88) 1,4-Dichlorobenzene	12.042	146	56297	17.956	ug/l	99
89) n-Butylbenzene	12.335	91	68297	15.517	ug/l	98
90) Hexachloroethane	12.536	117	15892	17.935	ug/l	93
91) 1,2-Dichlorobenzene	12.335	146	59023	19.863	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.945	75	10416	20.057	ug/l	84
93) 1,2,4-Trichlorobenzene	13.585	180	31843	18.314	ug/l	97
94) Hexachlorobutadiene	13.725	225	10466	15.581	ug/l	97
95) Naphthalene	13.774	128	130463	17.130	ug/l	99
96) 1,2,3-Trichlorobenzene	13.963	180	33281	19.266	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037360.D
Acq On : 30 Aug 2023 01:01
Operator : JC/MD
Sample : VX0829WBSD02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :


Quant Time: Aug 30 02:12:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Quant Time: Aug 30 02:12:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 94152
Ion Ratio Lower Upper
168 100
99 64.6 56.6 85.0



#2
Dichlorodifluoromethane
Concen: 16.512 ug/l
RT: 1.166 min Scan# 13
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 85 Resp: 18059
Ion Ratio Lower Upper
85 100
87 30.5 16.7 50.1








#3
Chloromethane
Concen: 24.641 ug/l
RT: 1.300 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :






Tgt Ion: 50 Resp: 23577
Ion Ratio Lower Upper
50 100
52 30.7 25.5 38.3




#4
Vinyl Chloride
Concen: 20.433 ug/l
RT: 1.374 min Scan# 47
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 62 Resp: 21281
Ion Ratio Lower Upper
62 100
64 33.2 25.4 38.0







Abundance Scan 84 (1.599 min): VX037061.D\data.ms (-78)



#5
Bromomethane
Concen: 19.010 ug/l
RT: 1.593 min Scan# 81
Delta R.T. -0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 9862
Ion Ratio Lower Upper
94 100
96 92.1 74.6 111.8



#6
Chloroethane
Concen: 22.445 ug/l
RT: 1.672 min Scan# 96
Delta R.T. 0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 64 Resp: 13756
Ion Ratio Lower Upper
64 100
66 34.3 25.1 37.7








#7
Trichlorofluoromethane
Concen: 17.291 ug/l
RT: 1.880 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:101 Resp: 29500
Ion Ratio Lower Upper
101 100
103 64.3 51.4 77.2



#8
Diethyl Ether
Concen: 24.889 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 74 Resp: 16428
Ion Ratio Lower Upper
74 100
45 83.0 53.3 159.8









#9
1,1,2-Trichlorotrifluoroethane
Concen: 15.396 ug/l
RT: 2.325 min Scan# 203
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:101 Resp: 15323
Ion Ratio Lower Upper
101 100
85 44.0 37.0 55.6
151 80.4 58.6 87.8



#10
Methyl Iodide
Concen: 17.801 ug/l
RT: 2.447 min Scan# 223
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion:142 Resp: 21533
Ion Ratio Lower Upper
142 100
127 46.1 41.6 62.4
141 13.6 11.6 17.4








#11
Tert butyl alcohol
Concen: 129.256 ug/l
RT: 2.977 min Scan# 310
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 59 Resp: 41156
Ion Ratio Lower Upper
59 100
57 9.3 6.0 9.0#



#12
1,1-Dichloroethene
Concen: 18.454 ug/l
RT: 2.312 min Scan# 201
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 96 Resp: 17886
Ion Ratio Lower Upper
96 100
61 151.4 144.6 216.8
98 63.7 50.4 75.6









#13
Acrolein
Concen: 70.892 ug/l
RT: 2.239 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 56 Resp: 17172
Ion Ratio Lower Upper
56 100
55 70.1 57.1 85.7



#14
Allyl chloride
Concen: 18.841 ug/l
RT: 2.660 min Scan# 258
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 41 Resp: 28264
Ion Ratio Lower Upper
41 100
39 63.6 55.8 83.6
76 38.6 24.3 36.5








#15
Acrylonitrile
Concen: 131.258 ug/l
RT: 3.068 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 53 Resp: 74596
Ion Ratio Lower Upper
53 100
52 81.7 66.0 99.0
51 35.0 29.8 44.6



#16
Acetone
Concen: 119.744 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 43 Resp: 63586
Ion Ratio Lower Upper
43 100
58 33.4 21.9 32.9#








#17
Carbon Disulfide
Concen: 17.128 ug/l
RT: 2.508 min Scan# 211
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 76 Resp: 42225
Ion Ratio Lower Upper
76 100
78 8.9 7.0 10.6



#18
Methyl Acetate
Concen: 24.556 ug/l
RT: 2.703 min Scan# 265
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 43 Resp: 49097
Ion Ratio Lower Upper
43 100
74 28.5 17.8 26.8#









#19
Methyl tert-butyl Ether
Concen: 23.469 ug/l
RT: 3.117 min Scan# 311
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 73 Resp: 86266
Ion Ratio Lower Upper
73 100
57 21.9 17.3 25.9



#20
Methylene Chloride
Concen: 23.974 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 84 Resp: 28469
Ion Ratio Lower Upper
84 100
49 106.6 103.6 155.4
51 35.6 33.8 50.6
86 66.0 49.4 74.0








#21
trans-1,2-Dichloroethene
Concen: 20.179 ug/l
RT: 3.087 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 96 Resp: 22232
Ion Ratio Lower Upper
96 100
61 127.1 119.2 178.8
98 65.9 48.7 73.1



#22
Diisopropyl ether
Concen: 23.215 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 45 Resp: 71464
Ion Ratio Lower Upper
45 100
43 74.8 69.7 104.5
87 32.8 20.8 31.2#
59 13.7 9.0 13.4#









#23
 Vinyl Acetate
 Concen: 115.407 ug/l
 RT: 3.721 min Scan# 413
 Delta R.T. -0.000 min
 Lab File: VX037360.D
 Acq: 30 Aug 2023 01:01




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 43 Resp: 272244
 Ion Ratio Lower Upper
 43 100
 86 12.7 7.8 11.8#



#24
 1,1-Dichloroethane
 Concen: 22.125 ug/l
 RT: 3.611 min Scan# 414
 Delta R.T. 0.006 min
 Lab File: VX037360.D
 Acq: 30 Aug 2023 01:01

Tgt Ion: 63 Resp: 42711
 Ion Ratio Lower Upper
 63 100
 98 7.8 3.3 9.8
 100 5.7 2.1 6.5





#25

2-Butanone

Concen: 123.721 ug/l

RT: 4.562 min Scan# 51

Delta R.T. 0.006 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 98921

Ion Ratio Lower Upper

43 100

72 29.8 19.4 29.2#



#26

2,2-Dichloropropane

Concen: 11.822 ug/l

RT: 4.471 min Scan# 555

Delta R.T. -0.006 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion: 77 Resp: 20381

Ion Ratio Lower Upper

77 100

97 25.5 10.8 32.3









#27
 cis-1,2-Dichloroethene
 Concen: 22.486 ug/l
 RT: 4.489 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: VX037360.D
 Acq: 30 Aug 2023 01:01

Instrument :
 MSVOA_X
 ClientSampleId :

Tgt Ion: 96 Resp: 29206




Ion	Ratio	Lower	Upper
96	100		
61	125.5	0.0	312.2
98	65.5	0.0	129.0




#28
 Bromochloromethane
 Concen: 25.544 ug/l
 RT: 4.891 min Scan# 624
 Delta R.T. -0.012 min
 Lab File: VX037360.D
 Acq: 30 Aug 2023 01:01

Tgt Ion: 49 Resp: 22028

Ion	Ratio	Lower	Upper
49	100		
129	2.0	0.0	3.8
130	96.1	57.5	86.3





#29

Tetrahydrofuran

Concen: 122.766 ug/l

RT: 5.007 min Scan# 643

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 42 Resp: 62598

Ion Ratio Lower Upper

42 100

72 54.4 34.2 51.4#

71 50.3 32.2 48.2#



#30

Chloroform

Concen: 22.333 ug/l

RT: 5.092 min Scan# 657

Delta R.T. -0.000 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion: 83 Resp: 48395

Ion Ratio Lower Upper

83 100


85 70.3 50.3 75.5








#31
Cyclohexane
Concen: 16.437 ug/l
RT: 5.470 min Scan# 719
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:	56	Resp:	25087
Ion	Ratio	Lower	Upper
56	100		
69	32.3	23.4	35.2
84	95.3	65.9	98.9




#32
1,1,1-Trichloroethane
Concen: 19.109 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01



Tgt Ion:	97	Resp:	37107
Ion	Ratio	Lower	Upper
97	100		
99	65.2	51.1	76.7
61	40.5	36.4	54.6





#33

1,2-Dichloroethane-d4

Concen: 65.202 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 65 Resp: 88503

Ion Ratio Lower Upper

65 100

67 52.0 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01




Tgt Ion: 114 Resp: 191512


Ion Ratio Lower Upper

114 100

63 19.1 0.0 45.0


88 16.4 0.0 36.6








#35
Dibromofluoromethane
Concen: 57.596 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:113 Resp: 73051
Ion Ratio Lower Upper
113 100
111 103.3 82.9 124.3
192 19.2 13.3 19.9



#36
1,1-Dichloropropene
Concen: 15.439 ug/l
RT: 5.696 min Scan# 756
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 75 Resp: 27821
Ion Ratio Lower Upper
75 100
110 35.4 17.0 50.9
77 30.0 24.7 37.1





#37

Ethyl Acetate

Concen: 20.132 ug/l

RT: 4.714 min Scan# 595

Delta R.T. -0.000 min


Lab File: VX037360.D

Acq: 30 Aug 2023 01:01




Instrument :

MSVOA_X

ClientSampleId :



Tgt Ion:	43	Resp:	37058
Ion	Ratio	Lower	Upper
43	100		
61	15.0	10.6	16.0
70	13.8	8.3	12.5



#38

Carbon Tetrachloride

Concen: 15.173 ug/l




RT: 5.678 min Scan# 753


Delta R.T. 0.006 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Tgt Ion:	117	Resp:	29700
Ion	Ratio	Lower	Upper
117	100		
119	93.2	78.2	117.2
121	29.6	24.4	36.6









#39
Methylcyclohexane
Concen: 13.888 ug/l
RT: 7.379 min Scan# 1032
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 26294
Ion Ratio Lower Upper
83 100
55 75.8 64.6 97.0
98 51.7 36.7 55.1



#40
Benzene
Concen: 19.101 ug/l
RT: 6.037 min Scan# 812
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 78 Resp: 98716
Ion Ratio Lower Upper
78 100
77 23.8 19.3 28.9








#41
Methacrylonitrile
Concen: 20.159 ug/l
RT: 4.922 min Scan# 61
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 41 Resp: 19290
Ion Ratio Lower Upper
41 100
39 55.2 45.2 67.8
67 89.4 52.6 78.8#
52 33.6 22.7 34.1



#42
1,2-Dichloroethane
Concen: 19.628 ug/l
RT: 6.086 min Scan# 820
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 62 Resp: 39779
Ion Ratio Lower Upper
62 100
98 10.0 0.0 16.0








#43
Isopropyl Acetate
Concen: 19.587 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 43 Resp: 57240
Ion Ratio Lower Upper
43 100
61 23.2 15.2 22.8#
87 16.7 9.8 14.6#



#44
Trichloroethene
Concen: 16.438 ug/l
RT: 7.129 min Scan# 991
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion:130 Resp: 24089
Ion Ratio Lower Upper
130 100
95 97.7 0.0 204.4





#45

1,2-Dichloropropane

Concen: 20.120 ug/l

RT: 7.433 min Scan# 1041

Delta R.T. 0.006 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 63 Resp: 25639

Ion Ratio Lower Upper

63 100

65 32.4 24.3 36.5



#46

Dibromomethane

Concen: 20.488 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01




Tgt Ion: 93 Resp: 20567


Ion Ratio Lower Upper

93 100

95 81.1 66.6 100.0

174 94.1 69.4 104.0









#47
Bromodichloromethane
Concen: 18.799 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 36146
Ion Ratio Lower Upper
83 100
85 62.0 49.8 74.6
127 9.0 7.1 10.7



#48
Methyl methacrylate
Concen: 19.753 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 41 Resp: 28152
Ion Ratio Lower Upper
41 100
69 95.1 59.2 88.8
39 57.5 47.7 71.5





#49

1,4-Dioxane

Concen: 455.677 ug/l

RT: 7.665 min Scan# 1078

Delta R.T. 0.006 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 88 Resp: 17490

Ion Ratio Lower Upper

88 100

43 27.9 0.0 0.0#

58 66.5 60.1 90.1



#50

Toluene-d8

Concen: 53.021 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.000 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion: 98 Resp: 243935

Ion Ratio Lower Upper

98 100

100 66.7 53.1 79.7





#51

4-Methyl-2-Pentanone

Concen: 108.132 ug/l

RT: 8.573 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 191885

Ion Ratio Lower Upper

43 100

58 44.2 30.4 45.6



#52

Toluene

Concen: 18.893 ug/l

RT: 8.720 min Scan# 1252

Delta R.T. -0.000 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion: 92 Resp: 64512

Ion Ratio Lower Upper

92 100

91 169.8 134.2 201.2









#53
t-1,3-Dichloropropene
Concen: 16.976 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 34754
Ion Ratio Lower Upper
75 100
77 33.0 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 17.633 ug/l
RT: 8.366 min Scan# 1194
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 75 Resp: 39012
Ion Ratio Lower Upper
75 100
77 31.2 25.3 37.9
39 40.9 46.7 70.1#









#55
1,1,2-Trichloroethane
Concen: 21.734 ug/l
RT: 9.153 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :

Tgt Ion: 97 Resp: 30220




Ion	Ratio	Lower	Upper
97	100		
83	81.8	65.8	98.8
85	54.9	41.6	62.4
99	63.9	47.8	71.8




#56
Ethyl methacrylate
Concen: 20.300 ug/l
RT: 9.116 min Scan# 1317
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 69 Resp: 44211

Ion	Ratio	Lower	Upper
69	100		
41	58.6	56.8	85.2
39	32.6	32.0	48.0









#57
1,3-Dichloropropane
Concen: 21.113 ug/l
RT: 9.311 min Scan# 11
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 76 Resp: 49488
Ion Ratio Lower Upper
76 100
78 32.7 26.1 39.1



#58
2-Chloroethyl Vinyl ether
Concen: 115.268 ug/l
RT: 8.244 min Scan# 1174
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 63 Resp: 120473
Ion Ratio Lower Upper
63 100
106 32.2 22.5 33.7





#59

2-Hexanone

Concen: 109.831 ug/l

RT: 9.433 min Scan# 1369

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 43 Resp: 148606

Ion Ratio Lower Upper

43 100

58 60.5 26.2 78.5



#60

Dibromochloromethane

Concen: 19.658 ug/l

RT: 9.518 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion: 129 Resp: 29101

Ion Ratio Lower Upper

129 100

127 78.5 38.8 116.4









#61
1,2-Dibromoethane
Concen: 20.880 ug/l
RT: 9.610 min Scan# 1398
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 31786
Ion Ratio Lower Upper
107 100
109 93.8 75.2 112.8



#62
4-Bromofluorobenzene
Concen: 56.361 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 95 Resp: 101903
Ion Ratio Lower Upper
95 100
174 73.7 0.0 135.6
176 72.2 0.0 131.0








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 185212
Ion Ratio Lower Upper
117 100
82 54.7 46.2 69.4
119 31.5 25.5 38.3




#64
Tetrachloroethene
Concen: 14.127 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01



Tgt Ion:164 Resp: 19087
Ion Ratio Lower Upper
164 100
166 134.1 102.7 154.1
129 107.3 80.4 120.6
131 100.4 76.5 114.7









#65
Chlorobenzene
Concen: 18.458 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 72253
Ion Ratio Lower Upper
112 100
114 31.5 25.7 38.5



#66
1,1,1,2-Tetrachloroethane
Concen: 18.668 ug/l
RT: 10.165 min Scan# 1489
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion:131 Resp: 26971
Ion Ratio Lower Upper
131 100
133 96.1 47.1 141.4
119 65.5 32.2 96.6





#67

Ethyl Benzene

Concen: 16.908 ug/l

RT: 10.195 min Scan# 1494

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 91 Resp: 116464

Ion Ratio Lower Upper

91 100

106 31.8 24.9 37.3



#68

m/p-Xylenes

Concen: 34.215 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion:106 Resp: 91012

Ion Ratio Lower Upper

106 100


91 205.0 165.5 248.3








#69
o-Xylene
Concen: 18.325 ug/l
RT: 10.640 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:106 Resp: 48091
Ion Ratio Lower Upper
106 100
91 215.6 108.9 326.8



#70
Styrene
Concen: 18.452 ug/l
RT: 10.658 min Scan# 1570
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion:104 Resp: 79770
Ion Ratio Lower Upper
104 100
78 51.0 43.6 65.4
103 54.3 43.6 65.4








#71
Bromoform
Concen: 19.115 ug/l
RT: 10.799 min Scan# 1593
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:173 Resp: 20658
Ion Ratio Lower Upper
173 100
175 48.7 24.1 72.4
254 0.0 0.0 0.0




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01



Tgt Ion:152 Resp: 91307
Ion Ratio Lower Upper
152 100
115 71.5 43.3 129.9
150 164.0 0.0 345.0









#73
Isopropylbenzene
Concen: 16.871 ug/l
RT: 10.963 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 111970
Ion Ratio Lower Upper
105 100
120 26.7 12.9 38.6



#74
N-ethyl acetate
Concen: 19.523 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 43 Resp: 50346
Ion Ratio Lower Upper
43 100
70 50.8 32.9 49.3#
55 27.9 19.3 28.9
61 27.6 18.8 28.2









#75
1,1,2,2-Tetrachloroethane
Concen: 21.567 ug/l
RT: 11.213 min Scan# 1661
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 83 Resp: 48203
Ion Ratio Lower Upper
83 100
131 10.5 5.3 15.8
85 64.6 32.0 96.2



#76
1,2,3-Trichloropropane
Concen: 25.138 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 75 Resp: 48882
Ion Ratio Lower Upper
75 100
77 32.9 19.1 57.5









#77
Bromobenzene
Concen: 19.170 ug/l
RT: 11.201 min Scan# 1659
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:156 Resp: 31071
Ion Ratio Lower Upper
156 100
77 163.9 88.4 265.3
158 98.9 48.0 144.0



#78
n-propylbenzene
Concen: 16.436 ug/l
RT: 11.305 min Scan# 1676
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 91 Resp: 121463
Ion Ratio Lower Upper
91 100
120 23.3 11.5 34.4









#79
2-Chlorotoluene
Concen: 18.170 ug/l
RT: 11.366 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 85808
Ion Ratio Lower Upper
91 100
126 33.5 16.3 48.8



#80
1,3,5-Trimethylbenzene
Concen: 17.573 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 105 Resp: 97288
Ion Ratio Lower Upper
105 100
120 49.2 24.1 72.3









#81
trans-1,4-Dichloro-2-butene
Concen: 14.555 ug/l
RT: 11.018 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 10133
Ion Ratio Lower Upper
75 100
53 104.0 90.5 135.7
89 45.8 36.9 55.3



#82
4-Chlorotoluene
Concen: 17.510 ug/l
RT: 11.457 min Scan# 1701
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion: 91 Resp: 95452
Ion Ratio Lower Upper
91 100
126 29.4 14.2 42.6









#83
tert-Butylbenzene
Concen: 18.431 ug/l
RT: 11.713 min Scan# 1743
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 97311
Ion Ratio Lower Upper
119 100
91 57.1 31.2 93.6
134 23.1 11.7 35.0



#84
1,2,4-Trimethylbenzene
Concen: 18.200 ug/l
RT: 11.750 min Scan# 1749
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion:105 Resp: 101831
Ion Ratio Lower Upper
105 100
120 43.8 21.2 63.6









#85
sec-Butylbenzene
Concen: 17.188 ug/l
RT: 11.890 min Scan# 1772
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 106000
Ion Ratio Lower Upper
105 100
134 20.0 9.8 29.4



#86
p-Isopropyltoluene
Concen: 17.220 ug/l
RT: 12.012 min Scan# 1792
Delta R.T. 0.006 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion:119 Resp: 91417
Ion Ratio Lower Upper
119 100
134 25.7 12.8 38.3
91 24.3 13.1 39.3





#87

1,3-Dichlorobenzene

Concen: 18.280 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:146 Resp: 55471

Ion Ratio Lower Upper

146 100

111 43.0 22.1 66.1

148 63.3 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 17.956 ug/l

RT: 12.042 min Scan# 1797

Delta R.T. -0.000 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01




Tgt Ion:146 Resp: 56297


Ion Ratio Lower Upper

146 100

111 43.4 21.3 63.7

148 63.2 31.8 95.4





#89

n-Butylbenzene

Concen: 15.517 ug/l

RT: 12.335 min Scan# 1845

Delta R.T. 0.006 min

Lab File: VX037360.D

Acq: 30 Aug 2023 01:01

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 91 Resp: 68297

Ion Ratio Lower Upper

91 100

92 53.3 27.4 82.0

134 25.7 12.6 37.8



#90

Hexachloroethane

Concen: 17.935 ug/l

RT: 12.536 min Scan# 1878

Delta R.T. -0.000 min

Lab File: VX037360.D




Acq: 30 Aug 2023 01:01


Tgt Ion: 117 Resp: 15892

Ion Ratio Lower Upper

117 100


201 71.3 32.9 98.6








#91
1,2-Dichlorobenzene
Concen: 19.863 ug/l
RT: 12.335 min Scan# 1845
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:146 Resp: 59023
Ion Ratio Lower Upper
146 100
111 43.5 22.4 67.0
148 64.1 31.6 95.0



#92
1,2-Dibromo-3-Chloropropane
Concen: 20.057 ug/l
RT: 12.945 min Scan# 1945
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Tgt Ion: 75 Resp: 10416
Ion Ratio Lower Upper
75 100
155 82.6 36.1 108.5
157 107.1 44.9 134.7








#93
1,2,4-Trichlorobenzene
Concen: 18.314 ug/l
RT: 13.585 min Scan# 2050
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:180 Resp: 31843
Ion Ratio Lower Upper
180 100
182 94.1 48.3 144.8
145 33.4 18.0 54.0




#94
Hexachlorobutadiene
Concen: 15.581 ug/l
RT: 13.725 min Scan# 2073
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01




Tgt Ion:225 Resp: 10466
Ion Ratio Lower Upper
225 100
223 64.4 31.1 93.2
227 63.2 32.5 97.5








#95
Naphthalene
Concen: 17.130 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Instrument :
MSVOA_X
ClientSampleId :



Tgt Ion:128 Resp: 130463
Ion Ratio Lower Upper
128 100
127 12.5 10.3 15.5
129 10.7 8.8 13.2



#96
1,2,3-Trichlorobenzene
Concen: 19.266 ug/l
RT: 13.963 min Scan# 2112
Delta R.T. -0.000 min
Lab File: VX037360.D
Acq: 30 Aug 2023 01:01

Tgt Ion:180 Resp: 33281
Ion Ratio Lower Upper
180 100
182 94.8 46.9 140.6
145 35.1 17.8 53.5

