

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037331.D
 Acq On : 29 Aug 2023 10:44
 Operator : JC/MD
 Sample : VX0829MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829MBL01

Quant Time: Aug 30 02:04:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	146459	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	238531	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95164	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87201	41.299	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.600%
35) Dibromofluoromethane	5.379	113	73096	46.271	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.540%
50) Toluene-d8	8.647	98	278061	48.525	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.040%
62) 4-Bromofluorobenzene	11.079	95	101916	45.257	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037331.D
Acq On : 29 Aug 2023 10:44
Operator : JC/MD
Sample : VX0829MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0829MBL01

Quant Time: Aug 30 02:04:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

