

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037335.D
 Acq On : 29 Aug 2023 12:25
 Operator : JC/MD
 Sample : 04126-12DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW305D-20230819DL

Quant Time: Aug 30 02:05:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	144138	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236562	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87605	42.158	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	84.320%		
35) Dibromofluoromethane	5.385	113	71265	45.487	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	90.980%		
50) Toluene-d8	8.647	98	276137	48.590	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.180%		
62) 4-Bromofluorobenzene	11.079	95	101292	45.355	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.700%		
Target Compounds						Qvalue
44) Trichloroethene	7.123	130	63052	34.833	ug/l	98
64) Tetrachloroethene	9.281	164	398	0.260	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037335.D
Acq On : 29 Aug 2023 12:25
Operator : JC/MD
Sample : 04126-12DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW305D-20230819DL

Quant Time: Aug 30 02:05:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

