

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037338.D
 Acq On : 29 Aug 2023 13:38
 Operator : JC/MD
 Sample : 04150-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB02-20230822

Quant Time: Aug 30 02:06:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145629	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	238864	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211933	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89321	42.544	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.080%
35) Dibromofluoromethane	5.385	113	72316	45.713	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.420%
50) Toluene-d8	8.647	98	278333	48.504	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.000%
62) 4-Bromofluorobenzene	11.079	95	101311	44.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.860%
Target Compounds						
30) Chloroform	5.093	83	4220	1.259	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037338.D
Acq On : 29 Aug 2023 13:38
Operator : JC/MD
Sample : 04150-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB02-20230822

Quant Time: Aug 30 02:06:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

