

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037341.D
 Acq On : 29 Aug 2023 14:51
 Operator : JC/MD
 Sample : 04207-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OW-07B-67-082823-H

Quant Time: Aug 30 02:07:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143289	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	233836	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206019	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86499	41.873	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	83.740%		
35) Dibromofluoromethane	5.385	113	72346	46.716	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	93.440%		
50) Toluene-d8	8.647	98	271678	48.363	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.720%		
62) 4-Bromofluorobenzene	11.079	95	99236	44.952	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	89.900%		
Target Compounds					Qvalue	
12) 1,1-Dichloroethene	2.313	96	86031	58.325	ug/l	84
16) Acetone	2.386	43	2148	2.658	ug/l	93
17) Carbon Disulfide	2.508	76	1316	0.351	ug/l	100
21) trans-1,2-Dichloroethene	3.093	96	3942	2.351	ug/l	91
24) 1,1-Dichloroethane	3.611	63	15732	5.355	ug/l	99
27) cis-1,2-Dichloroethene	4.489	96	107841	54.555	ug/l	79
42) 1,2-Dichloroethane	6.092	62	9349	3.778	ug/l	95
44) Trichloroethene	7.123	130	37674	21.055	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037341.D
Acq On : 29 Aug 2023 14:51
Operator : JC/MD
Sample : 04207-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OW-07B-67-082823-H

Quant Time: Aug 30 02:07:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

