

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037354.D
 Acq On : 29 Aug 2023 20:31
 Operator : JC/MD
 Sample : 04143-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 50122

Quant Time: Aug 30 02:10:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	247625	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	215112	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91797	42.096	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.200%
35) Dibromofluoromethane	5.385	113	75171	45.837	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.680%
50) Toluene-d8	8.647	98	287279	48.292	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.580%
62) 4-Bromofluorobenzene	11.079	95	104226	44.583	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.160%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20799	24.380	ug/l #	86
25) 2-Butanone	4.580	43	1429	1.112	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037354.D
Acq On : 29 Aug 2023 20:31
Operator : JC/MD
Sample : 04143-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
50122

Quant Time: Aug 30 02:10:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

