

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037364.D
 Acq On : 30 Aug 2023 02:45
 Operator : JC/MD
 Sample : 04196-25
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-345-VOC-04

Quant Time: Aug 30 03:49:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154461	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	253433	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	230207	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	95356	42.821	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.640%
35) Dibromofluoromethane	5.385	113	77950	46.442	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.880%
50) Toluene-d8	8.647	98	302107	49.621	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	112656	47.085	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.160%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	22013	25.269	ug/l	90
20) Methylene Chloride	2.788	84	15931	8.177	ug/l	86
25) 2-Butanone	4.568	43	7552	5.757	ug/l	95
43) Isopropyl Acetate	6.342	43	86315	22.320	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037364.D
Acq On : 30 Aug 2023 02:45
Operator : JC/MD
Sample : 04196-25
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-345-VOC-04

Quant Time: Aug 30 03:49:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

