

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037372.D
 Acq On : 30 Aug 2023 05:59
 Operator : JC/MD
 Sample : 04198-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OR-309-OK-370-VOC-26

Quant Time: Aug 31 01:26:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	150403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	251117	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	92502	42.661	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.320%
35) Dibromofluoromethane	5.385	113	76547	46.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.060%
50) Toluene-d8	8.647	98	298269	49.442	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.880%
62) 4-Bromofluorobenzene	11.079	95	109115	46.026	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.060%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	18756	22.111	ug/l #	86
20) Methylene Chloride	2.788	84	24434	12.880	ug/l #	84
37) Ethyl Acetate	4.733	43	2616	1.084	ug/l #	79
43) Isopropyl Acetate	6.342	43	90559	23.633	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037372.D
Acq On : 30 Aug 2023 05:59
Operator : JC/MD
Sample : 04198-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OR-309-OK-370-VOC-26

Quant Time: Aug 31 01:26:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

