

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037330.D
 Acq On : 29 Aug 2023 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 30 02:04:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.581	0.638	-9.8	93	0.00
3 P	Chloromethane	0.508	0.621	-22.2	96	0.00
4 C	Vinyl Chloride	0.553	0.625	-13.0#	95	0.00
5 T	Bromomethane	0.276	0.336	-21.7	106	0.00
6 T	Chloroethane	0.325	0.340	-4.6	92	0.00
7 T	Trichlorofluoromethane	0.906	0.978	-7.9	94	0.00
8 T	Diethyl Ether	0.351	0.400	-14.0	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.529	0.572	-8.1	97	0.00
10 T	Methyl Iodide	0.642	0.740	-15.3	89	0.00
11 T	Tert butyl alcohol	0.169	0.170	-0.6	81	-0.01
12 CM	1,1-Dichloroethene	0.515	0.546	-6.0#	91	0.00
13 T	Acrolein	0.129	0.112	13.2	70	0.00
14 T	Allyl chloride	0.797	0.863	-8.3	84	0.00
15 T	Acrylonitrile	0.302	0.343	-13.6	88	0.00
16 T	Acetone	0.282	0.325	-15.2	94	0.00
17 T	Carbon Disulfide	1.309	1.332	-1.8	89	0.00
18 T	Methyl Acetate	1.062	1.134	-6.8	83	0.00
19 T	Methyl tert-butyl Ether	1.952	2.159	-10.6	86	0.00
20 T	Methylene Chloride	0.631	0.687	-8.9	93	0.00
21 T	trans-1,2-Dichloroethene	0.585	0.618	-5.6	86	0.00
22 T	Diisopropyl ether	1.635	1.788	-9.4	86	0.00
23 T	Vinyl Acetate	1.253	1.386	-10.6	84	0.00
24 P	1,1-Dichloroethane	1.025	1.112	-8.5	87	0.00
25 T	2-Butanone	0.425	0.463	-8.9	85	0.00
26 T	2,2-Dichloropropane	0.916	0.971	-6.0	88	0.00
27 T	cis-1,2-Dichloroethene	0.690	0.759	-10.0	88	0.00
28 T	Bromochloromethane	0.458	0.495	-8.1	84	-0.01
29 T	Tetrahydrofuran	0.271	0.287	-5.9	82	0.00
30 C	Chloroform	1.151	1.244	-8.1#	86	0.00
31 T	Cyclohexane	0.811	0.883	-8.9	100	0.00
32 T	1,1,1-Trichloroethane	1.031	1.076	-4.4	86	0.00
33 S	1,2-Dichloroethane-d4	0.721	0.705	2.2	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	81	0.00
35 S	Dibromofluoromethane	0.331	0.351	-6.0	83	0.00
36 T	1,1-Dichloropropene	0.470	0.496	-5.5	90	0.00
37 T	Ethyl Acetate	0.481	0.506	-5.2	81	0.00
38 T	Carbon Tetrachloride	0.511	0.546	-6.8	89	0.00
39 T	Methylcyclohexane	0.494	0.628	-27.1#	120	0.00
40 TM	Benzene	1.349	1.527	-13.2	89	0.00
41 T	Methacrylonitrile	0.250	0.268	-7.2	80	0.00
42 TM	1,2-Dichloroethane	0.529	0.570	-7.8	82	0.00
43 T	Isopropyl Acetate	0.763	0.833	-9.2	82	0.00
44 TM	Trichloroethene	0.383	0.404	-5.5	86	0.00
45 C	1,2-Dichloropropane	0.333	0.381	-14.4#	88	0.00
46 T	Dibromomethane	0.262	0.296	-13.0	85	0.00
47 T	Bromodichloromethane	0.502	0.567	-12.9	85	0.00
48 T	Methyl methacrylate	0.372	0.406	-9.1	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
 Data File : VX037330.D
 Acq On : 29 Aug 2023 10:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 30 02:04:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.012	-20.0	88	0.00
50 S	Toluene-d8	1.201	1.229	-2.3	81	0.00
51 T	4-Methyl-2-Pentanone	0.463	0.516	-11.4	85	0.00
52 CM	Toluene	0.891	1.010	-13.4#	90	0.00
53 T	t-1,3-Dichloropropene	0.535	0.618	-15.5	85	0.00
54 T	cis-1,3-Dichloropropene	0.578	0.668	-15.6	87	0.00
55 T	1,1,2-Trichloroethane	0.363	0.425	-17.1	90	0.00
56 T	Ethyl methacrylate	0.569	0.660	-16.0	87	0.00
57 T	1,3-Dichloropropane	0.612	0.710	-16.0	90	0.00
58 T	2-Chloroethyl Vinyl ether	0.273	0.305	-11.7	85	0.00
59 T	2-Hexanone	0.353	0.406	-15.0	87	0.00
60 T	Dibromochloromethane	0.386	0.465	-20.5	88	0.00
61 T	1,2-Dibromoethane	0.397	0.461	-16.1	89	0.00
62 S	4-Bromofluorobenzene	0.472	0.512	-8.5	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	83	0.00
64 T	Tetrachloroethene	0.365	0.348	4.7	83	0.00
65 PM	Chlorobenzene	1.057	1.162	-9.9	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.390	0.443	-13.6	89	0.00
67 C	Ethyl Benzene	1.860	2.044	-9.9#	92	0.00
68 T	m/p-Xylenes	0.718	0.812	-13.1	94	0.00
69 T	o-Xylene	0.708	0.806	-13.8	93	0.00
70 T	Styrene	1.167	1.357	-16.3	93	0.00
71 P	Bromoform	0.292	0.355	-21.6	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	89	0.00
73 T	Isopropylbenzene	3.634	3.860	-6.2	98	0.00
74 T	N-amyl acetate	1.412	1.483	-5.0	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.224	1.371	-12.0	98	0.00
76 T	1,2,3-Trichloropropane	1.065	1.125	-5.6	93	0.00
77 T	Bromobenzene	0.888	0.958	-7.9	95	0.00
78 T	n-propylbenzene	4.047	4.442	-9.8	102	0.00
79 T	2-Chlorotoluene	2.586	2.772	-7.2	97	0.00
80 T	1,3,5-Trimethylbenzene	3.032	3.408	-12.4	102	0.00
81 T	trans-1,4-Dichloro-2-butene	0.381	0.423	-11.0	92	0.00
82 T	4-Chlorotoluene	2.985	3.199	-7.2	96	0.00
83 T	tert-Butylbenzene	2.891	3.379	-16.9	110	0.00
84 T	1,2,4-Trimethylbenzene	3.064	3.439	-12.2	102	0.00
85 T	sec-Butylbenzene	3.377	4.127	-22.2	117	0.00
86 T	p-Isopropyltoluene	2.907	3.566	-22.7	116	0.00
87 T	1,3-Dichlorobenzene	1.662	1.847	-11.1	100	0.00
88 T	1,4-Dichlorobenzene	1.717	1.858	-8.2	98	0.00
89 T	n-Butylbenzene	2.410	3.005	-24.7	120	0.00
90 T	Hexachloroethane	0.485	0.621	-28.0#	114	0.00
91 T	1,2-Dichlorobenzene	1.627	1.867	-14.8	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.284	0.319	-12.3	90	0.00
93 T	1,2,4-Trichlorobenzene	0.952	1.166	-22.5	111	0.00
94 T	Hexachlorobutadiene	0.368	0.494	-34.2#	135	0.00
95 T	Naphthalene	4.171	4.167	0.1	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082923\
Data File : VX037330.D
Acq On : 29 Aug 2023 10:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 30 02:04:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.946	1.152	-21.8	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6