

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
 Data File : VX047561.D
 Acq On : 29 Aug 2025 11:29
 Operator : JC/MD
 Sample : VX0829WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01

Quant Time: Aug 29 23:51:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	238310	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	414972	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372504	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	184959	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	172139	47.038	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.080%
35) Dibromofluoromethane	5.391	113	136803	48.033	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.060%
50) Toluene-d8	8.647	98	482880	48.370	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.740%
62) 4-Bromofluorobenzene	11.079	95	184241	49.126	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.260%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	47115	18.053	ug/l	99
3) Chloromethane	1.313	50	47681	16.728	ug/l	93
4) Vinyl Chloride	1.392	62	58806	17.551	ug/l	100
5) Bromomethane	1.630	94	28951	19.977	ug/l	93
6) Chloroethane	1.709	64	39850	18.719	ug/l	94
7) Trichlorofluoromethane	1.904	101	88022	18.470	ug/l	94
8) Diethyl Ether	2.148	74	34320	18.150	ug/l	96
9) 1,1,2-Trichlorotrifluo...	2.349	101	52657	18.553	ug/l	98
10) Methyl Iodide	2.471	142	77701	16.539	ug/l	97
11) Tert butyl alcohol	2.959	59	28236	83.943	ug/l	100
12) 1,1-Dichloroethene	2.337	96	53512	17.781	ug/l	96
13) Acrolein	2.252	56	54315	80.551	ug/l	96
14) Allyl chloride	2.678	41	98637	17.891	ug/l	98
15) Acrylonitrile	3.081	53	133241	88.521	ug/l	98
16) Acetone	2.392	43	119513	82.459	ug/l	98
17) Carbon Disulfide	2.526	76	157999	17.656	ug/l	99
18) Methyl Acetate	2.715	43	63366	17.733	ug/l	99
19) Methyl tert-butyl Ether	3.123	73	177564	18.076	ug/l	97
20) Methylene Chloride	2.800	84	64216	17.412	ug/l	99
21) trans-1,2-Dichloroethene	3.105	96	58493	18.264	ug/l	93
22) Diisopropyl ether	3.770	45	208150	18.746	ug/l	89
23) Vinyl Acetate	3.733	43	829387	91.729	ug/l	100
24) 1,1-Dichloroethane	3.623	63	111636	18.093	ug/l	99
25) 2-Butanone	4.562	43	167735	87.939	ug/l	95
26) 2,2-Dichloropropane	4.489	77	82028	18.393	ug/l	98
27) cis-1,2-Dichloroethene	4.501	96	70075	17.958	ug/l	99
28) Bromochloromethane	4.910	49	49766	18.691	ug/l	98
29) Tetrahydrofuran	5.019	42	101984	86.159	ug/l	99
30) Chloroform	5.105	83	113884	18.126	ug/l	97
31) Cyclohexane	5.489	56	96098	18.675	ug/l	97
32) 1,1,1-Trichloroethane	5.391	97	94442	18.450	ug/l	99
36) 1,1-Dichloropropene	5.702	75	78252	19.306	ug/l	99
37) Ethyl Acetate	4.727	43	78027	19.531	ug/l	98
38) Carbon Tetrachloride	5.684	117	82746	18.398	ug/l	98
39) Methylcyclohexane	7.379	83	92040	19.690	ug/l	96
40) Benzene	6.044	78	238732	18.449	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
 Data File : VX047561.D
 Acq On : 29 Aug 2025 11:29
 Operator : JC/MD
 Sample : VX0829WBSD01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01

Quant Time: Aug 29 23:51:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	4.934	41	38376	20.283	ug/l	98
42) 1,2-Dichloroethane	6.092	62	85532	18.361	ug/l	99
43) Isopropyl Acetate	6.342	43	120057	18.702	ug/l	99
44) Trichloroethene	7.129	130	59784	18.646	ug/l	95
45) 1,2-Dichloropropane	7.434	63	59261	18.181	ug/l	98
46) Dibromomethane	7.580	93	44266	18.448	ug/l	98
47) Bromodichloromethane	7.824	83	91210	18.587	ug/l	99
48) Methyl methacrylate	7.696	41	60616	18.356	ug/l	98
49) 1,4-Dioxane	7.683	88	12516	383.149	ug/l	95
51) 4-Methyl-2-Pentanone	8.574	43	353373	93.167	ug/l	99
52) Toluene	8.720	92	149613	18.972	ug/l	99
53) t-1,3-Dichloropropene	8.976	75	81987	18.545	ug/l	95
54) cis-1,3-Dichloropropene	8.366	75	92134	18.809	ug/l	99
55) 1,1,2-Trichloroethane	9.153	97	57431	18.668	ug/l	99
56) Ethyl methacrylate	9.116	69	85658	18.809	ug/l	100
57) 1,3-Dichloropropane	9.305	76	100569	18.881	ug/l	99
58) 2-Chloroethyl Vinyl ether	8.238	63	199294	102.330	ug/l	99
59) 2-Hexanone	9.427	43	237487	91.656	ug/l	99
60) Dibromochloromethane	9.519	129	68812	18.723	ug/l	99
61) 1,2-Dibromoethane	9.610	107	60047	19.180	ug/l	100
64) Tetrachloroethene	9.275	164	49533	19.848	ug/l	94
65) Chlorobenzene	10.079	112	167168	19.170	ug/l	99
66) 1,1,1,2-Tetrachloroethane	10.159	131	55640	18.816	ug/l	98
67) Ethyl Benzene	10.195	91	284675	19.427	ug/l	99
68) m/p-Xylenes	10.299	106	219405	40.313	ug/l	98
69) o-Xylene	10.640	106	103690	19.777	ug/l	99
70) Styrene	10.652	104	180297	19.819	ug/l	99
71) Bromoform	10.799	173	43704	19.410	ug/l #	100
73) Isopropylbenzene	10.957	105	271297	19.934	ug/l	99
74) N-amyl acetate	10.841	43	104394	18.574	ug/l	100
75) 1,1,2,2-Tetrachloroethane	11.207	83	82106	19.084	ug/l	100
76) 1,2,3-Trichloropropane	11.238	75	75262	22.799	ug/l	91
77) Bromobenzene	11.195	156	67488	19.384	ug/l	98
78) n-propylbenzene	11.299	91	321157	19.820	ug/l	100
79) 2-Chlorotoluene	11.360	91	196605	19.564	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	227135	20.411	ug/l	99
81) trans-1,4-Dichloro-2-b...	11.018	75	11826	19.507	ug/l	95
82) 4-Chlorotoluene	11.451	91	232287	19.700	ug/l	100
83) tert-Butylbenzene	11.713	119	222420	19.827	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	228228	20.120	ug/l	99
85) sec-Butylbenzene	11.890	105	275685	20.424	ug/l	99
86) p-Isopropyltoluene	12.006	119	229492	20.283	ug/l	99
87) 1,3-Dichlorobenzene	11.969	146	127668	19.668	ug/l	98
88) 1,4-Dichlorobenzene	12.036	146	130699	19.501	ug/l	100
89) n-Butylbenzene	12.329	91	213086	20.405	ug/l	99
90) Hexachloroethane	12.536	117	39843	19.157	ug/l	99
91) 1,2-Dichlorobenzene	12.335	146	122097	19.195	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	12.939	75	14719	18.548	ug/l	97
93) 1,2,4-Trichlorobenzene	13.585	180	78076	20.137	ug/l	99
94) Hexachlorobutadiene	13.725	225	27630	20.502	ug/l	98
95) Naphthalene	13.774	128	221986	19.586	ug/l	99
96) 1,2,3-Trichlorobenzene	13.957	180	73079	20.065	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
Data File : VX047561.D
Acq On : 29 Aug 2025 11:29
Operator : JC/MD
Sample : VX0829WBSD01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Quant Time: Aug 29 23:51:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Aug 29 23:51:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.556 min Scan# 71
Delta R.T. -0.012 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:168 Resp: 238310
Ion Ratio Lower Upper
168 100
99 61.0 48.9 73.3



#2
Dichlorodifluoromethane
Concen: 18.053 ug/l
RT: 1.185 min Scan# 16
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 85 Resp: 47115
Ion Ratio Lower Upper
85 100
87 31.4 15.9 47.7





#3

Chloromethane

Concen: 16.728 ug/l

RT: 1.313 min Scan# 31

Delta R.T. 0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 50 Resp: 47681

Ion Ratio Lower Upper

50 100

52 34.8 24.6 37.0



#4

Vinyl Chloride

Concen: 17.551 ug/l

RT: 1.392 min Scan# 50

Delta R.T. -0.000 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 62 Resp: 58806

Ion Ratio Lower Upper

62 100

64 32.2 25.9 38.9









#5
Bromomethane
Concen: 19.977 ug/l
RT: 1.630 min Scan# 89
Delta R.T. 0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 94 Resp: 28951
Ion Ratio Lower Upper
94 100
96 90.2 77.7 116.5



#6
Chloroethane
Concen: 18.719 ug/l
RT: 1.709 min Scan# 102
Delta R.T. 0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 64 Resp: 39850
Ion Ratio Lower Upper
64 100
66 30.7 27.4 41.2









#7
 Trichlorofluoromethane
 Concen: 18.470 ug/l
 RT: 1.904 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01


Tgt Ion:101 Resp: 88022
 Ion Ratio Lower Upper
 101 100
 103 67.1 50.1 75.1



#8
 Diethyl Ether
 Concen: 18.150 ug/l
 RT: 2.148 min Scan# 174
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29

Tgt Ion: 74 Resp: 34320
 Ion Ratio Lower Upper
 74 100
 45 101.6 49.0 147.0









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.553 ug/l
RT: 2.349 min Scan# 207
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:101 Resp: 52657
Ion Ratio Lower Upper
101 100
85 44.9 35.1 52.7
151 74.6 58.6 88.0



#10
Methyl Iodide
Concen: 16.539 ug/l
RT: 2.471 min Scan# 227
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion:142 Resp: 77701
Ion Ratio Lower Upper
142 100
127 44.6 37.6 56.4
141 13.2 10.6 16.0





#11

Tert butyl alcohol

Concen: 83.943 ug/l

RT: 2.959 min Scan# 308

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 59 Resp: 28236

Ion Ratio Lower Upper

59 100

57 9.9 7.9 11.9



#12

1,1-Dichloroethene

Concen: 17.781 ug/l

RT: 2.337 min Scan# 205

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion: 96 Resp: 53512


Ion Ratio Lower Upper

96 100

61 163.7 136.8 205.2

98 64.7 52.4 78.6









#13
Acrolein
Concen: 80.551 ug/l
RT: 2.252 min Scan# 191
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 56 Resp: 54315
Ion Ratio Lower Upper
56 100
55 68.0 57.0 85.6



#14
Allyl chloride
Concen: 17.891 ug/l
RT: 2.678 min Scan# 261
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 41 Resp: 98637
Ion Ratio Lower Upper
41 100
39 65.2 53.2 79.8
76 32.4 26.8 40.2









#15
Acrylonitrile
Concen: 88.521 ug/l
RT: 3.081 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 53 Resp: 133241
Ion Ratio Lower Upper
53 100
52 83.8 66.2 99.4
51 35.2 29.4 44.0



#16
Acetone
Concen: 82.459 ug/l
RT: 2.392 min Scan# 214
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 43 Resp: 119513
Ion Ratio Lower Upper
43 100
58 31.5 24.2 36.2









#17
Carbon Disulfide
Concen: 17.656 ug/l
RT: 2.526 min Scan# 21
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 76 Resp: 157999
Ion Ratio Lower Upper
76 100
78 8.8 7.4 11.0



#18
Methyl Acetate
Concen: 17.733 ug/l
RT: 2.715 min Scan# 267
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 43 Resp: 63366
Ion Ratio Lower Upper
43 100
74 22.9 17.8 26.8









#19
Methyl tert-butyl Ether
Concen: 18.076 ug/l
RT: 3.123 min Scan# 31
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 73 Resp: 177564
Ion Ratio Lower Upper
73 100
57 23.2 17.4 26.2



#20
Methylene Chloride
Concen: 17.412 ug/l
RT: 2.800 min Scan# 281
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 84 Resp: 64216
Ion Ratio Lower Upper
84 100
49 126.2 101.4 152.2
51 38.2 30.6 46.0
86 61.8 51.5 77.3









#21
trans-1,2-Dichloroethene
Concen: 18.264 ug/l
RT: 3.105 min Scan# 31
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 96 Resp: 58493
Ion Ratio Lower Upper
96 100
61 146.2 108.0 162.0
98 63.4 51.6 77.4



#22
Diisopropyl ether
Concen: 18.746 ug/l
RT: 3.770 min Scan# 440
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 45 Resp: 208150
Ion Ratio Lower Upper
45 100
43 118.6 83.2 124.8
87 27.1 20.2 30.4
59 11.4 9.4 14.0









#23
 Vinyl Acetate
 Concen: 91.729 ug/l
 RT: 3.733 min Scan# 41
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01


Tgt Ion: 43 Resp: 829387
 Ion Ratio Lower Upper
 43 100
 86 10.2 8.1 12.1



#24
 1,1-Dichloroethane
 Concen: 18.093 ug/l
 RT: 3.623 min Scan# 416
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29

Tgt Ion: 63 Resp: 111636
 Ion Ratio Lower Upper
 63 100
 98 7.4 3.8 11.2
 100 4.6 2.6 7.8





#25

2-Butanone

Concen: 87.939 ug/l

RT: 4.562 min Scan# 5

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 43 Resp: 167735

Ion Ratio Lower Upper

43 100

72 26.9 19.5 29.3



#26

2,2-Dichloropropane

Concen: 18.393 ug/l

RT: 4.489 min Scan# 558

Delta R.T. -0.006 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 77 Resp: 82028

Ion Ratio Lower Upper

77 100

97 22.2 11.6 34.8





#27

cis-1,2-Dichloroethene

Concen: 17.958 ug/l

RT: 4.501 min Scan# 51

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion: 96 Resp: 70075

Ion Ratio Lower Upper

96 100

61 146.0 0.0 289.4

98 64.7 0.0 129.2



#28

Bromochloromethane

Concen: 18.691 ug/l

RT: 4.910 min Scan# 627

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion: 49 Resp: 49766


Ion Ratio Lower Upper

49 100

129 1.7 0.0 1.8

130 76.4 59.8 89.6





#29

Tetrahydrofuran

Concen: 86.159 ug/l

RT: 5.019 min Scan# 64

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion: 42 Resp: 101984

Ion Ratio Lower Upper

42 100

72 43.5 34.5 51.7

71 39.2 31.7 47.5



#30

Chloroform

Concen: 18.126 ug/l

RT: 5.105 min Scan# 659

Delta R.T. -0.006 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 83 Resp: 113884

Ion Ratio Lower Upper

83 100

85 63.4 52.3 78.5









#31
Cyclohexane
Concen: 18.675 ug/l
RT: 5.489 min Scan# 71
Delta R.T. 0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 56 Resp: 96098
Ion Ratio Lower Upper
56 100
69 27.1 24.2 36.2
84 83.6 65.5 98.3



#32
1,1,1-Trichloroethane
Concen: 18.450 ug/l
RT: 5.391 min Scan# 706
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 97 Resp: 94442
Ion Ratio Lower Upper
97 100
99 63.7 51.5 77.3
61 46.8 36.5 54.7





#33

1,2-Dichloroethane-d4

Concen: 47.038 ug/l

RT: 5.958 min Scan# 799

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 65 Resp: 172139

Ion Ratio Lower Upper

65 100

67 53.6 0.0 104.8



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion: 114 Resp: 414972


Ion Ratio Lower Upper

114 100

63 22.0 0.0 43.6

88 16.7 0.0 33.2





#35

Dibromofluoromethane

Concen: 48.033 ug/l

RT: 5.391 min Scan# 707

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion:113 Resp: 136803

Ion Ratio Lower Upper

113 100

111 103.0 81.2 121.8

192 17.9 14.2 21.4



#36

1,1-Dichloropropene

Concen: 19.306 ug/l

RT: 5.702 min Scan# 757

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion: 75 Resp: 78252


Ion Ratio Lower Upper

75 100

110 35.1 17.1 51.3

77 30.8 25.0 37.6





#37

Ethyl Acetate

Concen: 19.531 ug/l

RT: 4.727 min Scan# 597

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion: 43 Resp: 78027

Ion Ratio Lower Upper

43 100

61 12.2 10.2 15.2

70 9.3 8.2 12.4



#38

Carbon Tetrachloride

Concen: 18.398 ug/l

RT: 5.684 min Scan# 754

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion: 117 Resp: 82746


Ion Ratio Lower Upper

117 100

119 93.6 76.4 114.6

121 29.1 24.9 37.3





#39

Methylcyclohexane

Concen: 19.690 ug/l

RT: 7.379 min Scan# 1

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion: 83 Resp: 92040

Ion Ratio Lower Upper

83 100

55 78.7 65.4 98.2

98 45.5 39.0 58.4



#40

Benzene

Concen: 18.449 ug/l

RT: 6.044 min Scan# 813

Delta R.T. -0.006 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 78 Resp: 238732

Ion Ratio Lower Upper

78 100

77 23.2 18.5 27.7









#41
Methacrylonitrile
Concen: 20.283 ug/l
RT: 4.934 min Scan# 611
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 41	Resp: 38376
Ion Ratio	Lower Upper
41 100	
39 54.2	42.4 63.6
67 72.9	60.6 90.8
52 30.2	23.8 35.8



#42
1,2-Dichloroethane
Concen: 18.361 ug/l
RT: 6.092 min Scan# 821
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 62	Resp: 85532
Ion Ratio	Lower Upper
62 100	
98 9.1	0.0 17.8









#43
Isopropyl Acetate
Concen: 18.702 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 43 Resp: 120057
Ion Ratio Lower Upper
43 100
61 19.5 15.4 23.2
87 11.6 9.9 14.9



#44
Trichloroethene
Concen: 18.646 ug/l
RT: 7.129 min Scan# 991
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion:130 Resp: 59784
Ion Ratio Lower Upper
130 100
95 98.6 0.0 206.6





#45

1,2-Dichloropropane

Concen: 18.181 ug/l

RT: 7.434 min Scan# 1041

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 63 Resp: 59261

Ion Ratio Lower Upper

63 100

65 32.3 24.7 37.1



#46

Dibromomethane

Concen: 18.448 ug/l

RT: 7.580 min Scan# 1065

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion: 93 Resp: 44266


Ion Ratio Lower Upper

93 100

95 82.4 66.7 100.1

174 96.0 75.2 112.8









#47
Bromodichloromethane
Concen: 18.587 ug/l
RT: 7.824 min Scan# 1105
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 83 Resp: 91210
Ion Ratio Lower Upper
83 100
85 65.4 52.0 78.0
127 9.6 6.8 10.2



#48
Methyl methacrylate
Concen: 18.356 ug/l
RT: 7.696 min Scan# 1084
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 41 Resp: 60616
Ion Ratio Lower Upper
41 100
69 81.3 63.4 95.2
39 57.9 44.7 67.1





#49

1,4-Dioxane

Concen: 383.149 ug/l

RT: 7.683 min Scan# 1081

Delta R.T. 0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion: 88 Resp: 12516

Ion Ratio Lower Upper

88 100

43 46.5 31.9 47.9

58 73.6 57.5 86.3



#50

Toluene-d8

Concen: 48.370 ug/l

RT: 8.647 min Scan# 1240

Delta R.T. -0.006 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 98 Resp: 482880

Ion Ratio Lower Upper

98 100

100 67.3 53.5 80.3









#51
4-Methyl-2-Pentanone
Concen: 93.167 ug/l
RT: 8.574 min Scan# 1228
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 43 Resp: 353373
Ion Ratio Lower Upper
43 100
58 37.5 30.4 45.6



#52
Toluene
Concen: 18.972 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 92 Resp: 149613
Ion Ratio Lower Upper
92 100
91 171.0 137.8 206.8









#53
 t-1,3-Dichloropropene
 Concen: 18.545 ug/l
 RT: 8.976 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01


Tgt Ion: 75 Resp: 81987
 Ion Ratio Lower Upper
 75 100
 77 29.4 25.7 38.5



#54
 cis-1,3-Dichloropropene
 Concen: 18.809 ug/l
 RT: 8.366 min Scan# 1194
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29

Tgt Ion: 75 Resp: 92134
 Ion Ratio Lower Upper
 75 100
 77 29.9 24.3 36.5
 39 47.9 39.0 58.4









#55
 1,1,2-Trichloroethane
 Concen: 18.668 ug/l
 RT: 9.153 min Scan# 1323
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01


Tgt Ion:	97	Resp:	57431
Ion	Ratio	Lower	Upper
97	100		
83	89.9	71.4	107.0
85	57.5	45.6	68.4
99	66.4	52.4	78.6



#56
 Ethyl methacrylate
 Concen: 18.809 ug/l
 RT: 9.116 min Scan# 1317
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29

Tgt Ion:	69	Resp:	85658
Ion	Ratio	Lower	Upper
69	100		
41	68.6	55.0	82.4
39	37.0	29.8	44.6





#57

1,3-Dichloropropane

Concen: 18.881 ug/l

RT: 9.305 min Scan# 11348

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 76 Resp: 100569

Ion Ratio Lower Upper

76 100

78 31.7 25.8 38.6



#58

2-Chloroethyl Vinyl ether

Concen: 102.330 ug/l

RT: 8.238 min Scan# 1173

Delta R.T. -0.007 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 63 Resp: 199294

Ion Ratio Lower Upper

63 100

106 27.3 22.2 33.4





#59

2-Hexanone

Concen: 91.656 ug/l

RT: 9.427 min Scan# 1368

Delta R.T. -0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 43 Resp: 237487

Ion Ratio Lower Upper

43 100

58 54.3 26.7 80.0



#60

Dibromochloromethane

Concen: 18.723 ug/l

RT: 9.519 min Scan# 1383

Delta R.T. -0.000 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 129 Resp: 68812

Ion Ratio Lower Upper

129 100

127 76.1 38.5 115.3









#61
1,2-Dibromoethane
Concen: 19.180 ug/l
RT: 9.610 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:107 Resp: 60047
Ion Ratio Lower Upper
107 100
109 94.3 75.5 113.3



#62
4-Bromofluorobenzene
Concen: 49.126 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 95 Resp: 184241
Ion Ratio Lower Upper
95 100
174 74.4 0.0 145.2
176 71.3 0.0 140.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:117 Resp: 372504
Ion Ratio Lower Upper
117 100
82 56.2 44.6 66.8
119 33.3 25.7 38.5



#64
Tetrachloroethene
Concen: 19.848 ug/l
RT: 9.275 min Scan# 1343
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion:164 Resp: 49533
Ion Ratio Lower Upper
164 100
166 121.6 102.1 153.1
129 93.3 79.3 118.9
131 89.6 77.1 115.7









#65
Chlorobenzene
Concen: 19.170 ug/l
RT: 10.079 min Scan# 1475
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:112 Resp: 167168
Ion Ratio Lower Upper
112 100
114 31.0 25.3 37.9



#66
1,1,1,2-Tetrachloroethane
Concen: 18.816 ug/l
RT: 10.159 min Scan# 1488
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion:131 Resp: 55640
Ion Ratio Lower Upper
131 100
133 96.6 47.4 142.3
119 66.6 32.6 97.7





#67

Ethyl Benzene

Concen: 19.427 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 91 Resp: 284675

Ion Ratio Lower Upper

91 100

106 30.2 23.9 35.9



#68

m/p-Xylenes

Concen: 40.313 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 106 Resp: 219405

Ion Ratio Lower Upper

106 100

91 205.0 166.1 249.1





#69

o-Xylene

Concen: 19.777 ug/l

RT: 10.640 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion:106 Resp: 103690

Ion Ratio Lower Upper

106 100

91 221.0 109.9 329.7



#70

Styrene

Concen: 19.819 ug/l

RT: 10.652 min Scan# 1569

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion:104 Resp: 180297


Ion Ratio Lower Upper

104 100

78 51.5 41.6 62.4

103 54.3 43.0 64.4





#71

Bromoform

Concen: 19.410 ug/l

RT: 10.799 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion:173 Resp: 43704

Ion Ratio Lower Upper

173 100

175 49.2 24.5 73.5

254 0.0 0.1 0.1#



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion:152 Resp: 184959


Ion Ratio Lower Upper

152 100

115 72.6 42.9 128.5

150 167.8 0.0 351.0









#73
Isopropylbenzene
Concen: 19.934 ug/l
RT: 10.957 min Scan# 1619
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:105 Resp: 271297
Ion Ratio Lower Upper
105 100
120 26.0 12.8 38.4



#74
N-ethyl acetate
Concen: 18.574 ug/l
RT: 10.841 min Scan# 1600
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 43 Resp: 104394
Ion Ratio Lower Upper
43 100
70 41.4 32.9 49.3
55 24.3 19.2 28.8
61 24.0 19.0 28.6









#75
1,1,2,2-Tetrachloroethane
Concen: 19.084 ug/l
RT: 11.207 min Scan# 1660
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 83 Resp: 82106
Ion Ratio Lower Upper
83 100
131 9.7 5.1 15.3
85 65.0 32.4 97.0



#76
1,2,3-Trichloropropane
Concen: 22.799 ug/l
RT: 11.238 min Scan# 1665
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 75 Resp: 75262
Ion Ratio Lower Upper
75 100
77 34.0 19.7 59.1









#77
Bromobenzene
Concen: 19.384 ug/l
RT: 11.195 min Scan# 1658
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion:156 Resp: 67488
Ion Ratio Lower Upper
156 100
77 171.2 84.4 253.1
158 97.3 47.4 142.3



#78
n-propylbenzene
Concen: 19.820 ug/l
RT: 11.299 min Scan# 1675
Delta R.T. -0.006 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 91 Resp: 321157
Ion Ratio Lower Upper
91 100
120 22.4 11.2 33.6





#79

2-Chlorotoluene

Concen: 19.564 ug/l

RT: 11.360 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :





VX0829WBSD01

Tgt Ion: 91 Resp: 196605

Ion Ratio Lower Upper

91 100

126 33.0 16.4 49.2



#80

1,3,5-Trimethylbenzene

Concen: 20.411 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion: 105 Resp: 227135

Ion Ratio Lower Upper

105 100

120 47.3 23.8 71.5









#81
trans-1,4-Dichloro-2-butene
Concen: 19.507 ug/l
RT: 11.018 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 75 Resp: 11826
Ion Ratio Lower Upper
75 100
53 112.4 85.7 128.5
89 47.9 36.4 54.6



#82
4-Chlorotoluene
Concen: 19.700 ug/l
RT: 11.451 min Scan# 1700
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 91 Resp: 232287
Ion Ratio Lower Upper
91 100
126 29.0 14.6 43.8





#83

tert-Butylbenzene

Concen: 19.827 ug/l

RT: 11.713 min Scan# 1743

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion:119 Resp: 222420

Ion Ratio Lower Upper

119 100

91 58.3 28.5 85.5

134 23.1 11.3 33.9



#84

1,2,4-Trimethylbenzene

Concen: 20.120 ug/l

RT: 11.750 min Scan# 1749

Delta R.T. -0.000 min

Lab File: VX047561.D




Acq: 29 Aug 2025 11:29


Tgt Ion:105 Resp: 228228

Ion Ratio Lower Upper

105 100

120 45.1 22.3 66.8









#85
 sec-Butylbenzene
 Concen: 20.424 ug/l
 RT: 11.890 min Scan# 1772
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01


Tgt Ion:105 Resp: 275685
 Ion Ratio Lower Upper
 105 100
 134 19.4 9.8 29.5



#86
 p-Isopropyltoluene
 Concen: 20.283 ug/l
 RT: 12.006 min Scan# 1791
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29

Tgt Ion:119 Resp: 229492
 Ion Ratio Lower Upper
 119 100
 134 25.4 13.0 38.9
 91 25.1 12.4 37.2





#87

1,3-Dichlorobenzene

Concen: 19.668 ug/l

RT: 11.969 min Scan# 1785

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion:146 Resp: 127668

Ion Ratio Lower Upper

146 100

111 42.5 20.6 61.9

148 62.5 31.9 95.5



#88

1,4-Dichlorobenzene

Concen: 19.501 ug/l

RT: 12.036 min Scan# 1796

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion:146 Resp: 130699


Ion Ratio Lower Upper

146 100

111 40.2 20.1 60.2

148 63.6 31.8 95.4









#89
n-Butylbenzene
Concen: 20.405 ug/l
RT: 12.329 min Scan# 1844
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29




Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01


Tgt Ion: 91 Resp: 213086
Ion Ratio Lower Upper
91 100
92 54.6 27.0 81.0
134 26.0 12.6 37.6



#90
Hexachloroethane
Concen: 19.157 ug/l
RT: 12.536 min Scan# 1878
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29

Tgt Ion: 117 Resp: 39843
Ion Ratio Lower Upper
117 100
201 66.9 33.7 101.1









#91
 1,2-Dichlorobenzene
 Concen: 19.195 ug/l
 RT: 12.335 min Scan# 1845
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29




Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829WBSD01


Tgt Ion:146 Resp: 122097
 Ion Ratio Lower Upper
 146 100
 111 43.1 21.7 65.1
 148 63.3 32.0 96.2



#92
 1,2-Dibromo-3-Chloropropane
 Concen: 18.548 ug/l
 RT: 12.939 min Scan# 1944
 Delta R.T. -0.000 min
 Lab File: VX047561.D
 Acq: 29 Aug 2025 11:29

Tgt Ion: 75 Resp: 14719
 Ion Ratio Lower Upper
 75 100
 155 79.6 41.9 125.6
 157 102.5 51.8 155.4





#93

1,2,4-Trichlorobenzene

Concen: 20.137 ug/l

RT: 13.585 min Scan# 2050

Delta R.T. -0.000 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29

Instrument :

MSVOA_X

ClientSampleId :

VX0829WBSD01





Tgt Ion:180 Resp: 78076

Ion Ratio Lower Upper

180 100

182 93.7 47.7 143.1

145 32.4 16.3 48.8



#94

Hexachlorobutadiene

Concen: 20.502 ug/l

RT: 13.725 min Scan# 2073

Delta R.T. 0.006 min

Lab File: VX047561.D

Acq: 29 Aug 2025 11:29




Tgt Ion:225 Resp: 27630


Ion Ratio Lower Upper

225 100

223 62.4 31.1 93.5


227 61.1 31.8 95.3








#95
Naphthalene
Concen: 19.586 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29


Instrument :
MSVOA_X
ClientSampleId :
VX0829WBSD01



Tgt Ion:128 Resp: 221986
Ion Ratio Lower Upper
128 100
127 12.9 10.0 15.0
129 11.0 8.7 13.1



#96
1,2,3-Trichlorobenzene
Concen: 20.065 ug/l
RT: 13.957 min Scan# 2111
Delta R.T. -0.000 min
Lab File: VX047561.D
Acq: 29 Aug 2025 11:29



Tgt Ion:180 Resp: 73079
Ion Ratio Lower Upper
180 100
182 94.3 48.1 144.4
145 33.5 17.6 52.9

