

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
 Data File : VX047558.D
 Acq On : 29 Aug 2025 10:14
 Operator : JC/MD
 Sample : VX0829MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0829MBL01

Quant Time: Aug 29 23:49:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	171489	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	6.763	114	340643	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	341218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	171442	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	139195	52.856	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.720%	
35) Dibromofluoromethane	5.391	113	117614	50.306	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.620%	
50) Toluene-d8	8.647	98	418768	51.101	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.200%	
62) 4-Bromofluorobenzene	11.079	95	174899	56.810	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 113.620%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
Data File : VX047558.D
Acq On : 29 Aug 2025 10:14
Operator : JC/MD
Sample : VX0829MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0829MBL01

Quant Time: Aug 29 23:49:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

