

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
 Data File : VX047564.D
 Acq On : 29 Aug 2025 12:42
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Aug 29 23:53:58 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	284261	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	570660	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	562526	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	283455	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	229282	52.524	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.040%	
35) Dibromofluoromethane	5.397	113	196503	50.171	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.340%	
50) Toluene-d8	8.653	98	697476	50.805	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.600%	
62) 4-Bromofluorobenzene	11.079	95	286515	55.553	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 111.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
Data File : VX047564.D
Acq On : 29 Aug 2025 12:42
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Aug 29 23:53:58 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

