

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
 Data File : VX047566.D
 Acq On : 29 Aug 2025 13:24
 Operator : JC/MD
 Sample : Q2987-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Aug 29 23:54:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 29 02:31:18 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	194659	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	390750	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	397369	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	200391	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	162922	54.502	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.000%
35) Dibromofluoromethane	5.398	113	134869	50.289	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.580%
50) Toluene-d8	8.653	98	483022	51.383	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.760%
62) 4-Bromofluorobenzene	11.079	95	202701	57.398	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	114.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX082925\
Data File : VX047566.D
Acq On : 29 Aug 2025 13:24
Operator : JC/MD
Sample : Q2987-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Aug 29 23:54:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X082825W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 29 02:31:18 2025
Response via : Initial Calibration

