

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083018\
 Data File : VX004318.D
 Acq On : 30 Aug 2018 13:09
 Operator : JC/MD
 Sample : VX0830WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0830WBL01

Quant Time: Aug 30 15:25:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	307313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	484936	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	452023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	251607	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	212331	53.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.02%	
35) Dibromofluoromethane	5.50	113	189668	52.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.14%	
50) Toluene-d8	8.71	98	689277	50.62	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.24%	
62) 4-Bromofluorobenzene	11.14	95	237651	46.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083018\
Data File : VX004318.D
Acq On : 30 Aug 2018 13:09
Operator : JC/MD
Sample : VX0830WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01

Quant Time: Aug 30 15:25:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

