

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083018\
 Data File : VX004323.D
 Acq On : 30 Aug 2018 17:19
 Operator : JC/MD
 Sample : J4735-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KG-BASELINE-WATER

Quant Time: Aug 31 16:27:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	240229	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	377803	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	347702	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193464	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	170831	54.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.12%	
35) Dibromofluoromethane	5.50	113	148153	52.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.42%	
50) Toluene-d8	8.71	98	534071	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	11.14	95	182709	46.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.32%	
Target Compounds						
19) Methyl tert-butyl Ether	3.20	73	4743	0.57	ug/l	96
68) m/p-Xylenes	10.36	106	2180	0.39	ug/l	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083018\
Data File : VX004323.D
Acq On : 30 Aug 2018 17:19
Operator : JC/MD
Sample : J4735-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KG-BASELINE-WATER

Quant Time: Aug 31 16:27:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

