

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083018\
 Data File : VX004326.D
 Acq On : 30 Aug 2018 18:39
 Operator : JC/MD
 Sample : J4282-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-01-082918-C

Quant Time: Aug 31 01:15:55 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	250452	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	407303	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	388153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	208324	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	181401	55.57	ug/l	0.00
Spiked Amount			Recovery	=	111.14%	
35) Dibromofluoromethane	5.50	113	158951	52.46	ug/l	0.00
Spiked Amount			Recovery	=	104.92%	
50) Toluene-d8	8.71	98	582665	50.94	ug/l	0.00
Spiked Amount			Recovery	=	101.88%	
62) 4-Bromofluorobenzene	11.14	95	203061	47.58	ug/l	0.00
Spiked Amount			Recovery	=	95.16%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Methyl Iodide	2.51	142	1792	4.78	ug/l	99
16) Acetone	2.45	43	18466	5.98	ug/l	100
18) Methyl Acetate	2.78	43	6535	1.35	ug/l	98
20) Methylene Chloride	2.85	84	1138106	362.19	ug/l	98
30) Chloroform	5.21	83	6743	1.27	ug/l	95
43) Isopropyl Acetate	6.46	43	195045	28.19	ug/l	99
95) Naphthalene	13.83	128	15851	1.17	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083018\
Data File : VX004326.D
Acq On : 30 Aug 2018 18:39
Operator : JC/MD
Sample : J4282-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-01-082918-C

Quant Time: Aug 31 01:15:55 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

