

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083018\
 Data File : VX004319.D
 Acq On : 30 Aug 2018 13:36
 Operator : JC/MD
 Sample : VX0830MBL01
 Misc : 5.00µ/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0830MBL01

Quant Time: Aug 30 15:28:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	260986	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	409489	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	362952	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213497	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	186315	54.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	109.54%	
35) Dibromofluoromethane	5.50	113	164019	53.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	107.68%	
50) Toluene-d8	8.71	98	591365	51.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.86%	
62) 4-Bromofluorobenzene	11.14	95	205865	47.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.96%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083018\
Data File : VX004319.D
Acq On : 30 Aug 2018 13:36
Operator : JC/MD
Sample : VX0830MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0830MBL01

Quant Time: Aug 30 15:28:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.66 min Scan# 832

Delta R.T. -0.00 min

Lab File: VX004319.D

Acq: 30 Aug 2018 13:36

Instrument :

MSVOA_X

ClientSampleId :


VX0830MBL01

Tot Ion:168 Resp: 260986




Ion Ratio Lower Upper

168 100

99 50.3 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): VX0



#33

1,2-Dichloroethane-d4

Concen: 54.77 ug/l

RT: 6.07 min Scan# 899

Delta R.T. -0.00 min

Lab File: VX004319.D


Acq: 30 Aug 2018 13:36

Tgt Ion: 65 Resp: 186315



Ion Ratio Lower Upper


65 100

67 53.8 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.86 min Scan# 1029
 Delta R.T. -0.00 min
 Lab File: VX004319.D
 Acq: 30 Aug 2018 13:36

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0830MBL01

Tot Ion:114 Resp: 409489
 Ion Ratio Lower Upper
 114 100
 63 19.7 0.0 39.0
 88 15.5 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 53.84 ug/l
 RT: 5.50 min Scan# 806
 Delta R.T. -0.00 min
 Lab File: VX004319.D
 Acq: 30 Aug 2018 13:36

Tgt Ion:113 Resp: 164019
 Ion Ratio Lower Upper
 113 100
 111 102.0 82.4 123.6
 192 23.6 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 51.43 ug/l
RT: 8.71 min Scan# 1333
Delta R.T. -0.00 min
Lab File: VX004319.D
Acq: 30 Aug 2018 13:36




Instrument :
MSVOA_X
ClientSampleId :
VX0830MBL01


Tot Ion: 98 Resp: 591365
Ion Ratio Lower Upper
98 100
100 65.3 52.9 79.3



#62
4-Bromofluorobenzene
Concen: 47.98 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX004319.D
Acq: 30 Aug 2018 13:36

Tgt Ion: 95 Resp: 205865
Ion Ratio Lower Upper
95 100
174 93.0 0.0 185.2
176 87.6 0.0 178.6









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX004319.D
Acq: 30 Aug 2018 13:36

Instrument :
MSVOA_X
ClientSampleId :
VX0830MBL01

Tot Ion:117 Resp: 362952
Ion Ratio Lower Upper
117 100
82 56.8 47.8 71.6
119 33.0 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.00 min
Lab File: VX004319.D
Acq: 30 Aug 2018 13:36

Tgt Ion:152 Resp: 213497
Ion Ratio Lower Upper
152 100
115 50.5 39.0 117.0
150 151.7 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

