

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083019\
 Data File : VX011932.D
 Acq On : 31 Aug 2019 03:44
 Operator : SY/SP
 Sample : K4493-05
 Misc : 5.34g/5.0mL/MSVOA X/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T9-12-13-I-(0)

Quant Time: Aug 31 05:10:02 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 01:59:05 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	434722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	575836	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	512942	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	273494	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	181984	43.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.38%	
35) Dibromofluoromethane	5.48	113	167184	45.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.56%	
50) Toluene-d8	8.71	98	653523	49.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.02%	
62) 4-Bromofluorobenzene	11.13	95	244037	42.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.30%	
Target Compounds						
16) Acetone	2.43	43	80972	90.356	ug/l	96
18) Methyl Acetate	2.76	43	53461	21.551	ug/l	97
29) Tetrahydrofuran	5.13	42	5012	6.152	ug/l #	84
52) Toluene	8.78	92	10562	1.172	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083019\
Data File : VX011932.D
Acq On : 31 Aug 2019 03:44
Operator : SY/SP
Sample : K4493-05
Misc : 5.34g/5.0mL/MSVOA X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T9-12-13-I-(0)

Quant Time: Aug 31 05:10:02 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 01:59:05 2019
Response via : Initial Calibration

