


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083019\
Data File : VX011900.D
Acq On : 30 Aug 2019 08:26
Operator : SY/SP
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_X/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082919S.M
Title : SW846 8260
Last Update : Fri Aug 30 01:59:05 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1627

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.2	18052	PASS
75	95	30	60	49.6	49248	PASS
95	95	100	100	100.0	99322	PASS
96	95	5	9	6.7	6648	PASS
173	174	0.00	2	1.0	999	PASS
174	95	50	100	98.7	98045	PASS
175	174	5	9	7.2	7054	PASS
176	174	95	101	97.8	95858	PASS
177	176	5	9	6.8	6511	PASS