

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024006.D
 Acq On : 30 Aug 2021 10:13
 Operator : JC/MD
 Sample : VX0830MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0830MBL01

Quant Time: Aug 31 03:09:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	184129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	313923	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	289189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	120051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	133720	52.034	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.060%
35) Dibromofluoromethane	5.391	113	103347	49.739	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.480%
50) Toluene-d8	8.653	98	382958	48.426	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.860%
62) 4-Bromofluorobenzene	11.085	95	138245	47.487	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024006.D
Acq On : 30 Aug 2021 10:13
Operator : JC/MD
Sample : VX0830MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0830MBL01

Quant Time: Aug 31 03:09:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

