

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024010.D
 Acq On : 30 Aug 2021 13:31
 Operator : JC/MD
 Sample : M3561-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUII-E

Quant Time: Aug 31 03:10:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	182792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	309174	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	289070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121692	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	132973	52.122	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.240%
35) Dibromofluoromethane	5.397	113	103158	50.411	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.820%
50) Toluene-d8	8.653	98	380130	48.807	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.620%
62) 4-Bromofluorobenzene	11.085	95	138440	48.285	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%
Target Compounds						
16) Acetone	2.392	43	2931	2.569	ug/l	Qvalue # 72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024010.D
Acq On : 30 Aug 2021 13:31
Operator : JC/MD
Sample : M3561-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUII-E

Quant Time: Aug 31 03:10:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

