

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024011.D
 Acq On : 30 Aug 2021 13:55
 Operator : JC/MD
 Sample : M3547-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK-8-25-21

Quant Time: Aug 31 03:11:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	161774	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	274389	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	251124	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102853	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	116522	51.607	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.220%	
35) Dibromofluoromethane	5.397	113	91379	50.316	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.640%	
50) Toluene-d8	8.653	98	334789	48.434	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.860%	
62) 4-Bromofluorobenzene	11.085	95	119649	47.021	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.040%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.800	84	3298	1.710	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024011.D
Acq On : 30 Aug 2021 13:55
Operator : JC/MD
Sample : M3547-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK-8-25-21

Quant Time: Aug 31 03:11:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

