

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024018.D
 Acq On : 30 Aug 2021 16:40
 Operator : JC/MD
 Sample : M3583-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GR-OF-20210827

Quant Time: Aug 31 03:12:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	160841	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	278525	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112984	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	119365	53.173	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	106.340%
35) Dibromofluoromethane	5.397	113	90529	49.108	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.220%
50) Toluene-d8	8.653	98	339050	48.322	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.640%
62) 4-Bromofluorobenzene	11.085	95	128052	49.576	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.160%
Target Compounds						Qvalue
16) Acetone	2.386	43	3223	3.211	ug/l	99
27) cis-1,2-Dichloroethene	4.507	96	2938	1.498	ug/l	94
44) Trichloroethene	7.129	130	5677	2.694	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024018.D
Acq On : 30 Aug 2021 16:40
Operator : JC/MD
Sample : M3583-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GR-OF-20210827

Quant Time: Aug 31 03:12:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

