

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024020.D
 Acq On : 30 Aug 2021 17:27
 Operator : JC/MD
 Sample : M3583-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-117T-20210827

Quant Time: Aug 31 03:12:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	165766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	280426	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	259332	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112252	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	120154	51.934	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.860%	
35) Dibromofluoromethane	5.397	113	92938	50.073	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.140%	
50) Toluene-d8	8.653	98	344873	48.819	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 97.640%	
62) 4-Bromofluorobenzene	11.085	95	125803	48.375	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.760%	
Target Compounds						
						Qvalue
16) Acetone	2.392	43	3276	3.167	ug/l #	84
27) cis-1,2-Dichloroethene	4.501	96	18202	9.002	ug/l	96
44) Trichloroethene	7.129	130	16268	7.669	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024020.D
Acq On : 30 Aug 2021 17:27
Operator : JC/MD
Sample : M3583-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-117T-20210827

Quant Time: Aug 31 03:12:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

