

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024027.D
 Acq On : 30 Aug 2021 20:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 31 03:13:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	88	0.00
2 T	Dichlorodifluoromethane	0.528	0.543	-2.8	83	0.00
3 P	Chloromethane	0.579	0.565	2.4	81	0.00
4 C	Vinyl Chloride	0.565	0.574	-1.6#	83	0.00
5 T	Bromomethane	0.355	0.325	8.5	78	0.00
6 T	Chloroethane	0.383	0.370	3.4	84	0.00
7 T	Trichlorofluoromethane	0.892	0.956	-7.2	88	0.00
8 T	Diethyl Ether	0.311	0.337	-8.4	88	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.502	0.544	-8.4	88	0.00
10 T	Methyl Iodide	0.699	0.738	-5.6	82	0.00
11 T	Tert butyl alcohol	0.141	0.169	-19.9	97	0.00
12 CM	1,1-Dichloroethene	0.480	0.533	-11.0#	91	0.00
13 T	Acrolein	0.084	0.092	-9.5	91	0.00
14 T	Allyl chloride	0.884	0.997	-12.8	92	0.00
15 T	Acrylonitrile	0.314	0.352	-12.1	91	0.00
16 T	Acetone	0.312	0.329	-5.4	90	0.00
17 T	Carbon Disulfide	1.144	1.298	-13.5	90	0.00
18 T	Methyl Acetate	0.779	0.876	-12.5	94	0.00
19 T	Methyl tert-butyl Ether	1.670	2.010	-20.4	97	0.00
20 T	Methylene Chloride	0.596	0.622	-4.4	91	0.00
21 T	trans-1,2-Dichloroethene	0.513	0.563	-9.7	90	0.00
22 T	Diisopropyl ether	1.761	2.065	-17.3	95	0.00
23 T	Vinyl Acetate	1.425	1.695	-18.9	94	0.00
24 P	1,1-Dichloroethane	0.980	1.140	-16.3	95	0.00
25 T	2-Butanone	0.473	0.527	-11.4	90	0.00
26 T	2,2-Dichloropropane	0.805	0.771	4.2	78	0.00
27 T	cis-1,2-Dichloroethene	0.610	0.683	-12.0	92	0.00
28 T	Bromochloromethane	0.446	0.466	-4.5	84	0.00
29 T	Tetrahydrofuran	0.306	0.343	-12.1	91	0.00
30 C	Chloroform	1.038	1.181	-13.8#	95	0.00
31 T	Cyclohexane	0.897	1.005	-12.0	92	0.00
32 T	1,1,1-Trichloroethane	0.889	1.066	-19.9	96	0.00
33 S	1,2-Dichloroethane-d4	0.698	0.696	0.3	85	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.331	0.320	3.3	82	0.00
36 T	1,1-Dichloropropene	0.463	0.528	-14.0	93	0.00
37 T	Ethyl Acetate	0.569	0.635	-11.6	95	0.00
38 T	Carbon Tetrachloride	0.483	0.570	-18.0	96	0.00
39 T	Methylcyclohexane	0.564	0.624	-10.6	89	0.00
40 TM	Benzene	1.369	1.503	-9.8	91	0.00
41 T	Methacrylonitrile	0.309	0.347	-12.3	93	0.00
42 TM	1,2-Dichloroethane	0.538	0.629	-16.9	96	0.00
43 T	Isopropyl Acetate	0.890	1.020	-14.6	95	0.00
44 TM	Trichloroethene	0.378	0.405	-7.1	90	0.00
45 C	1,2-Dichloropropane	0.364	0.399	-9.6#	90	0.00
46 T	Dibromomethane	0.253	0.285	-12.6	93	0.00
47 T	Bromodichloromethane	0.458	0.554	-21.0	96	0.00
48 T	Methyl methacrylate	0.434	0.510	-17.5	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024027.D
 Acq On : 30 Aug 2021 20:12
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 31 03:13:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	91	0.00
50 S	Toluene-d8	1.260	1.184	6.0	80	0.00
51 T	4-Methyl-2-Pentanone	0.576	0.675	-17.2	93	0.00
52 CM	Toluene	0.887	0.986	-11.2#	91	0.00
53 T	t-1,3-Dichloropropene	0.521	0.618	-18.6	93	0.00
54 T	cis-1,3-Dichloropropene	0.555	0.637	-14.8	92	0.00
55 T	1,1,2-Trichloroethane	0.365	0.411	-12.6	93	0.00
56 T	Ethyl methacrylate	0.569	0.686	-20.6	95	0.00
57 T	1,3-Dichloropropane	0.606	0.689	-13.7	92	0.00
58 T	2-Chloroethyl Vinyl ether	0.294	0.306	-4.1	86	0.00
59 T	2-Hexanone	0.442	0.536	-21.3	95	0.00
60 T	Dibromochloromethane	0.353	0.433	-22.7	94	0.00
61 T	1,2-Dibromoethane	0.386	0.444	-15.0	91	0.00
62 S	4-Bromofluorobenzene	0.464	0.489	-5.4	87	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.399	0.414	-3.8	88	0.00
65 PM	Chlorobenzene	1.043	1.159	-11.1	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.434	-17.0	94	0.00
67 C	Ethyl Benzene	1.822	2.077	-14.0#	92	0.00
68 T	m/p-Xylenes	0.694	0.802	-15.6	93	0.00
69 T	o-Xylene	0.682	0.798	-17.0	94	0.00
70 T	Styrene	1.111	1.346	-21.2	95	0.00
71 P	Bromoform	0.263	0.342	-30.0#	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.814	4.241	-11.2	94	0.00
74 T	N-amyl acetate	1.724	1.998	-15.9	97	0.00
75 P	1,1,2,2-Tetrachloroethane	1.250	1.356	-8.5	97	0.00
76 T	1,2,3-Trichloropropane	1.166	1.294	-11.0	96	0.00
77 T	Bromobenzene	0.939	1.041	-10.9	95	0.00
78 T	n-propylbenzene	4.362	4.885	-12.0	94	0.00
79 T	2-Chlorotoluene	2.636	3.003	-13.9	96	0.00
80 T	1,3,5-Trimethylbenzene	3.134	3.604	-15.0	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.429	-8.1	90	0.00
82 T	4-Chlorotoluene	2.993	3.436	-14.8	95	0.00
83 T	tert-Butylbenzene	3.044	3.527	-15.9	97	0.00
84 T	1,2,4-Trimethylbenzene	3.125	3.579	-14.5	96	0.00
85 T	sec-Butylbenzene	3.854	4.457	-15.6	94	0.00
86 T	p-Isopropyltoluene	3.172	3.699	-16.6	94	0.00
87 T	1,3-Dichlorobenzene	1.699	1.871	-10.1	95	0.00
88 T	1,4-Dichlorobenzene	1.695	1.880	-10.9	95	0.00
89 T	n-Butylbenzene	2.832	3.215	-13.5	92	0.00
90 T	Hexachloroethane	0.488	0.622	-27.5#	99	0.00
91 T	1,2-Dichlorobenzene	1.646	1.810	-10.0	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.257	0.316	-23.0	100	0.00
93 T	1,2,4-Trichlorobenzene	0.991	1.076	-8.6	92	0.00
94 T	Hexachlorobutadiene	0.442	0.478	-8.1	95	0.00
95 T	Naphthalene	3.231	3.797	-17.5	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024027.D
Acq On : 30 Aug 2021 20:12
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 31 03:13:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.987	1.115	-13.0	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6