

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
 Data File : VX024007.D
 Acq On : 30 Aug 2021 10:37
 Operator : JC/MD
 Sample : VX0830WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0830WBL01

Quant Time: Aug 31 03:10:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 17 11:57:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	167150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	282280	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258095	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103934	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	120299	51.567	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 103.140%	
35) Dibromofluoromethane	5.397	113	93493	50.041	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.080%	
50) Toluene-d8	8.653	98	344457	48.440	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.880%	
62) 4-Bromofluorobenzene	11.085	95	122672	46.862	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 93.720%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083021\
Data File : VX024007.D
Acq On : 30 Aug 2021 10:37
Operator : JC/MD
Sample : VX0830WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01

Quant Time: Aug 31 03:10:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081721W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 17 11:57:51 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. 0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37




Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01


Tgt Ion:168 Resp: 167150
Ion Ratio Lower Upper
168 100
99 63.1 45.8 68.6



#33
1,2-Dichloroethane-d4
Concen: 51.567 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37

Tgt Ion: 65 Resp: 120299
Ion Ratio Lower Upper
65 100
67 51.1 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37




Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01


Tgt Ion:114 Resp: 282280
Ion Ratio Lower Upper
114 100
63 22.4 0.0 40.8
88 17.3 0.0 32.8



#35
Dibromofluoromethane
Concen: 50.041 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37

Tgt Ion:113 Resp: 93493
Ion Ratio Lower Upper
113 100
111 101.6 81.9 122.9
192 19.0 16.7 25.1









#50
Toluene-d8
Concen: 48.440 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37




Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01


Tgt Ion: 98 Resp: 344457
Ion Ratio Lower Upper
98 100
100 64.2 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 46.862 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37

Tgt Ion: 95 Resp: 122672
Ion Ratio Lower Upper
95 100
174 77.6 0.0 154.6
176 75.2 0.0 155.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.006 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37

Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01

Tgt Ion:117 Resp: 258095
Ion Ratio Lower Upper
117 100
82 60.3 47.4 71.2
119 32.1 26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VX024007.D
Acq: 30 Aug 2021 10:37

Tgt Ion:152 Resp: 103934
Ion Ratio Lower Upper
152 100
115 60.2 41.1 123.3
150 154.8 0.0 349.0

