

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030942.D
 Acq On : 30 Aug 2022 16:35
 Operator : JC/MD
 Sample : VX0830WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0830WBL01

Quant Time: Aug 30 17:05:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	207773	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	349101	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	309960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	132528	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152775	51.266	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	102.540%
35) Dibromofluoromethane	5.391	113	134299	52.188	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.380%
50) Toluene-d8	8.653	98	509591	53.471	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.940%
62) 4-Bromofluorobenzene	11.085	95	166914	49.183	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030942.D
Acq On : 30 Aug 2022 16:35
Operator : JC/MD
Sample : VX0830WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0830WBL01

Quant Time: Aug 30 17:05:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

