

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030946.D
 Acq On : 30 Aug 2022 18:09
 Operator : JC/MD
 Sample : N4371-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A4127

Quant Time: Aug 30 19:02:01 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	217897	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	367321	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	319765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130792	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	127673	40.852	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	81.700%
35) Dibromofluoromethane	5.391	113	113541	41.933	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	83.860%
50) Toluene-d8	8.653	98	423521	42.236	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	84.480%#
62) 4-Bromofluorobenzene	11.085	95	135374	37.911	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	75.820%#
Target Compounds						
					Qvalue	
10) Methyl Iodide	2.465	142	1713	5.004	ug/l #	93
11) Tert butyl alcohol	2.971	59	5072	6.850	ug/l	97
16) Acetone	2.386	43	3440	2.439	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030946.D
Acq On : 30 Aug 2022 18:09
Operator : JC/MD
Sample : N4371-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4127

Quant Time: Aug 30 19:02:01 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

