

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030949.D
 Acq On : 30 Aug 2022 19:20
 Operator : JC/MD
 Sample : N4281-05DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE115D1-20220818DL

Quant Time: Aug 31 03:15:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	215260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	356215	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321216	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137802	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	148614	48.135	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.260%		
35) Dibromofluoromethane	5.391	113	131879	50.224	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	100.440%		
50) Toluene-d8	8.653	98	506847	52.121	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	104.240%		
62) 4-Bromofluorobenzene	11.085	95	168740	48.728	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.460%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.166	85	854	0.375	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.325	101	6859	2.586	ug/l	98
12) 1,1-Dichloroethene	2.319	96	1222	0.467	ug/l	95
27) cis-1,2-Dichloroethene	4.501	96	1582	0.484	ug/l #	20
30) Chloroform	5.105	83	1815	0.342	ug/l #	79
38) Carbon Tetrachloride	5.678	117	1784	0.480	ug/l #	78
44) Trichloroethene	7.135	130	145189	43.814	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030949.D
Acq On : 30 Aug 2022 19:20
Operator : JC/MD
Sample : N4281-05DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE115D1-20220818DL

Quant Time: Aug 31 03:15:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

