

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030950.D
 Acq On : 30 Aug 2022 19:44
 Operator : JC/MD
 Sample : N4281-09DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D4-20220818DL

Quant Time: Aug 31 03:15:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	204760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	345883	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	310958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	140762	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147798	50.325	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 100.660%	
35) Dibromofluoromethane	5.391	113	129375	50.742	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.480%	
50) Toluene-d8	8.653	98	497486	52.687	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.380%	
62) 4-Bromofluorobenzene	11.085	95	161021	47.888	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.780%	
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	1968	0.780	ug/l	96
27) cis-1,2-Dichloroethene	4.483	96	1093	0.352	ug/l	92
44) Trichloroethene	7.129	130	103897	32.290	ug/l	97
64) Tetrachloroethene	9.275	164	3168	1.077	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030950.D
Acq On : 30 Aug 2022 19:44
Operator : JC/MD
Sample : N4281-09DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D4-20220818DL

Quant Time: Aug 31 03:15:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

