

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030953.D
 Acq On : 30 Aug 2022 20:54
 Operator : JC/MD
 Sample : N4297-04DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW305D-20220820DL

Quant Time: Aug 31 03:16:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	202259	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	362017	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	305837	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	135028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	147420	50.817	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.640%
35) Dibromofluoromethane	5.391	113	126368	47.354	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.700%
50) Toluene-d8	8.653	98	477923	48.359	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.720%
62) 4-Bromofluorobenzene	11.085	95	158882	45.146	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.300%
Target Compounds						
20) Methylene Chloride	2.788	84	1173	0.394	ug/l	93
44) Trichloroethene	7.129	130	124317	36.914	ug/l	95
64) Tetrachloroethene	9.281	164	692	0.239	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030953.D
Acq On : 30 Aug 2022 20:54
Operator : JC/MD
Sample : N4297-04DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW305D-20220820DL

Quant Time: Aug 31 03:16:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

