

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX083022\  
 Data File : VX030955.D  
 Acq On : 30 Aug 2022 21:41  
 Operator : JC/MD  
 Sample : N4297-15DL 10X  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 29 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE131D3-20220821DL

Quant Time: Aug 31 03:17:18 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X083022W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 30 14:24:53 2022  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|------------------------------|--------|-------|----------|----------|--------|----------|
| -----                        |        |       |          |          |        |          |
| Internal Standards           |        |       |          |          |        |          |
| 1) Pentafluorobenzene        | 5.562  | 168   | 195357   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.763  | 114   | 339573   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 283019   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 117310   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964  | 65    | 138452   | 49.412   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 74 - 125 | Recovery | =      | 98.820%  |
| 35) Dibromofluoromethane     | 5.391  | 113   | 121140   | 48.395   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =      | 96.800%  |
| 50) Toluene-d8               | 8.653  | 98    | 453547   | 48.926   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 86 - 113 | Recovery | =      | 97.860%  |
| 62) 4-Bromofluorobenzene     | 11.085 | 95    | 146533   | 44.389   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | =      | 88.780%  |
| Target Compounds             |        |       |          |          |        |          |
|                              |        |       |          |          | Qvalue |          |
| 9) 1,1,2-Trichlorotrifluo... | 2.331  | 101   | 72139    | 29.972   | ug/l   | 96       |
| 12) 1,1-Dichloroethene       | 2.319  | 96    | 604      | 0.254    | ug/l # | 63       |
| 20) Methylene Chloride       | 2.788  | 84    | 1172     | 0.408    | ug/l   | 91       |
| 44) Trichloroethene          | 7.135  | 130   | 9281     | 2.938    | ug/l   | 96       |
| 64) Tetrachloroethene        | 9.275  | 164   | 3253     | 1.215    | ug/l   | 93       |
| -----                        |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX083022\  
Data File : VX030955.D  
Acq On : 30 Aug 2022 21:41  
Operator : JC/MD  
Sample : N4297-15DL 10X  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 29 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE131D3-20220821DL

Quant Time: Aug 31 03:17:18 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X083022W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 30 14:24:53 2022  
Response via : Initial Calibration

