

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX083022\  
 Data File : VX030956.D  
 Acq On : 30 Aug 2022 22:05  
 Operator : JC/MD  
 Sample : N4281-11  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 30 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 RE132D6-20220818

Quant Time: Aug 31 03:17:38 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X083022W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Aug 30 14:24:53 2022  
 Response via : Initial Calibration

| Compound                     | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|------------------------------|--------|-------|----------|----------|--------|----------|
| -----                        |        |       |          |          |        |          |
| Internal Standards           |        |       |          |          |        |          |
| 1) Pentafluorobenzene        | 5.562  | 168   | 199641   | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.769  | 114   | 333217   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.055 | 117   | 297964   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.024 | 152   | 117484   | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds  |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4    | 5.964  | 65    | 141467   | 49.405   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 74 - 125 | Recovery | =      | 98.800%  |
| 35) Dibromofluoromethane     | 5.391  | 113   | 127108   | 51.748   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 75 - 124 | Recovery | =      | 103.500% |
| 50) Toluene-d8               | 8.653  | 98    | 470529   | 51.726   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 86 - 113 | Recovery | =      | 103.460% |
| 62) 4-Bromofluorobenzene     | 11.085 | 95    | 151754   | 46.847   | ug/l   | 0.00     |
| Spiked Amount                | 50.000 | Range | 83 - 123 | Recovery | =      | 93.700%  |
| Target Compounds             |        |       |          |          |        |          |
|                              |        |       |          |          | Qvalue |          |
| 9) 1,1,2-Trichlorotrifluo... | 2.325  | 101   | 1600     | 0.651    | ug/l   | 98       |
| 44) Trichloroethene          | 7.129  | 130   | 85963    | 27.731   | ug/l   | 98       |
| -----                        |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX083022\  
Data File : VX030956.D  
Acq On : 30 Aug 2022 22:05  
Operator : JC/MD  
Sample : N4281-11  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 30 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
RE132D6-20220818

Quant Time: Aug 31 03:17:38 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X083022W.M  
Quant Title : SW846 8260  
QLast Update : Tue Aug 30 14:24:53 2022  
Response via : Initial Calibration

