

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030963.D
 Acq On : 31 Aug 2022 00:50
 Operator : JC/MD
 Sample : N4375-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-4

Quant Time: Aug 31 03:20:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	210142	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	370967	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	326722	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	145536	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	150678	49.992	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.980%
35) Dibromofluoromethane	5.397	113	136441	49.895	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.780%
50) Toluene-d8	8.653	98	498704	49.244	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.480%
62) 4-Bromofluorobenzene	11.085	95	163857	45.436	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.880%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	33969	24.971	ug/l	98
20) Methylene Chloride	2.794	84	10972	3.548	ug/l #	89
37) Ethyl Acetate	4.727	43	31045	7.266	ug/l	98
43) Isopropyl Acetate	6.342	43	126508	19.080	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030963.D
Acq On : 31 Aug 2022 00:50
Operator : JC/MD
Sample : N4375-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-4

Quant Time: Aug 31 03:20:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

