

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030941.D
 Acq On : 30 Aug 2022 16:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 30 17:07:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	91	0.00
2 T	Dichlorodifluoromethane	0.529	0.517	2.3	89	0.00
3 P	Chloromethane	0.533	0.474	11.1	87	0.00
4 C	Vinyl Chloride	0.630	0.564	10.5#	87	0.00
5 T	Bromomethane	0.235	0.194	17.4	90	0.00
6 T	Chloroethane	0.379	0.334	11.9	88	0.00
7 T	Trichlorofluoromethane	0.958	0.865	9.7	88	0.00
8 T	Diethyl Ether	0.370	0.358	3.2	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.616	0.575	6.7	92	0.00
10 T	Methyl Iodide	0.465	0.468	-0.6	78	0.00
11 T	Tert butyl alcohol	0.170	0.181	-6.5	103	-0.01
12 CM	1,1-Dichloroethene	0.608	0.556	8.6#	89	0.00
13 T	Acrolein	0.078	0.087	-11.5	97	0.00
14 T	Allyl chloride	0.986	0.939	4.8	90	0.00
15 T	Acrylonitrile	0.375	0.400	-6.7	100	0.00
16 T	Acetone	0.324	0.324	0.0	100	0.00
17 T	Carbon Disulfide	1.604	1.430	10.8	88	0.00
18 T	Methyl Acetate	0.949	0.950	-0.1	99	0.00
19 T	Methyl tert-butyl Ether	1.979	2.027	-2.4	97	0.00
20 T	Methylene Chloride	0.736	0.687	6.7	97	0.00
21 T	trans-1,2-Dichloroethene	0.659	0.618	6.2	91	0.00
22 T	Diisopropyl ether	2.031	1.966	3.2	94	0.00
23 T	Vinyl Acetate	1.630	1.705	-4.6	97	0.00
24 P	1,1-Dichloroethane	1.190	1.138	4.4	92	0.00
25 T	2-Butanone	0.524	0.545	-4.0	99	0.00
26 T	2,2-Dichloropropane	0.896	0.749	16.4	80	0.00
27 T	cis-1,2-Dichloroethene	0.759	0.737	2.9	93	0.00
28 T	Bromochloromethane	0.385	0.383	0.5	103	0.00
29 T	Tetrahydrofuran	0.339	0.356	-5.0	100	0.00
30 C	Chloroform	1.234	1.195	3.2#	95	0.00
31 T	Cyclohexane	1.061	0.992	6.5	90	0.00
32 T	1,1,1-Trichloroethane	1.046	0.966	7.6	88	0.00
33 S	1,2-Dichloroethane-d4	0.717	0.743	-3.6	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	94	0.00
35 S	Dibromofluoromethane	0.369	0.376	-1.9	102	0.00
36 T	1,1-Dichloropropene	0.548	0.484	11.7	89	0.00
37 T	Ethyl Acetate	0.576	0.590	-2.4	100	0.00
38 T	Carbon Tetrachloride	0.522	0.489	6.3	90	0.00
39 T	Methylcyclohexane	0.670	0.617	7.9	88	0.00
40 TM	Benzene	1.579	1.486	5.9	92	0.00
41 T	Methacrylonitrile	0.313	0.315	-0.6	96	0.00
42 TM	1,2-Dichloroethane	0.524	0.519	1.0	95	0.00
43 T	Isopropyl Acetate	0.894	0.891	0.3	96	0.00
44 TM	Trichloroethene	0.465	0.422	9.2	93	0.00
45 C	1,2-Dichloropropane	0.421	0.400	5.0#	93	0.00
46 T	Dibromomethane	0.288	0.283	1.7	96	0.00
47 T	Bromodichloromethane	0.553	0.535	3.3	91	0.00
48 T	Methyl methacrylate	0.447	0.448	-0.2	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
 Data File : VX030941.D
 Acq On : 30 Aug 2022 16:03
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 30 17:07:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 30 14:24:53 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.012	0.013	-8.3	99	-0.01
50 S	Toluene-d8	1.365	1.367	-0.1	98	0.00
51 T	4-Methyl-2-Pentanone	0.590	0.599	-1.5	93	0.00
52 CM	Toluene	0.965	0.945	2.1#	93	0.00
53 T	t-1,3-Dichloropropene	0.558	0.577	-3.4	94	0.00
54 T	cis-1,3-Dichloropropene	0.639	0.637	0.3	92	0.00
55 T	1,1,2-Trichloroethane	0.418	0.424	-1.4	99	0.00
56 T	Ethyl methacrylate	0.613	0.653	-6.5	98	0.00
57 T	1,3-Dichloropropane	0.690	0.689	0.1	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.308	0.338	-9.7	101	-0.01
59 T	2-Hexanone	0.439	0.466	-6.2	99	0.00
60 T	Dibromochloromethane	0.416	0.433	-4.1	97	0.00
61 T	1,2-Dibromoethane	0.432	0.450	-4.2	97	0.00
62 S	4-Bromofluorobenzene	0.486	0.502	-3.3	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.473	0.435	8.0	91	0.00
65 PM	Chlorobenzene	1.182	1.117	5.5	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.419	0.410	2.1	94	0.00
67 C	Ethyl Benzene	2.097	2.030	3.2#	94	0.00
68 T	m/p-Xylenes	0.794	0.773	2.6	92	0.00
69 T	o-Xylene	0.777	0.773	0.5	92	0.00
70 T	Styrene	1.312	1.291	1.6	91	0.00
71 P	Bromoform	0.329	0.344	-4.6	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.956	3.752	5.2	92	0.00
74 T	N-amyl acetate	1.529	1.634	-6.9	102	0.00
75 P	1,1,2,2-Tetrachloroethane	1.355	1.350	0.4	100	0.00
76 T	1,2,3-Trichloropropane	1.231	1.261	-2.4	100	0.00
77 T	Bromobenzene	0.959	0.937	2.3	97	0.00
78 T	n-propylbenzene	4.636	4.418	4.7	90	0.00
79 T	2-Chlorotoluene	2.879	2.648	8.0	90	0.00
80 T	1,3,5-Trimethylbenzene	3.333	3.258	2.3	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.414	0.414	0.0	92	0.00
82 T	4-Chlorotoluene	3.219	3.052	5.2	93	0.00
83 T	tert-Butylbenzene	3.272	3.176	2.9	91	0.00
84 T	1,2,4-Trimethylbenzene	3.412	3.449	-1.1	91	0.00
85 T	sec-Butylbenzene	4.247	4.098	3.5	90	0.00
86 T	p-Isopropyltoluene	3.523	3.465	1.6	94	0.00
87 T	1,3-Dichlorobenzene	1.890	1.797	4.9	90	0.00
88 T	1,4-Dichlorobenzene	1.938	1.892	2.4	96	0.00
89 T	n-Butylbenzene	3.298	3.240	1.8	90	0.00
90 T	Hexachloroethane	0.619	0.615	0.6	93	0.00
91 T	1,2-Dichlorobenzene	1.917	1.890	1.4	94	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.338	0.354	-4.7	94	0.00
93 T	1,2,4-Trichlorobenzene	1.226	1.246	-1.6	94	0.00
94 T	Hexachlorobutadiene	0.520	0.509	2.1	94	0.00
95 T	Naphthalene	4.383	4.674	-6.6	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083022\
Data File : VX030941.D
Acq On : 30 Aug 2022 16:03
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 30 17:07:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X083022W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 30 14:24:53 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.237	1.256	-1.5	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6