

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037385.D
 Acq On : 30 Aug 2023 13:39
 Operator : JC/MD
 Sample : 04150-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D1-20230822

Quant Time: Aug 31 02:22:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	234829	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207257	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92574	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86792	42.075	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.160%
35) Dibromofluoromethane	5.385	113	71187	45.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.540%
50) Toluene-d8	8.647	98	274031	48.575	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.160%
62) 4-Bromofluorobenzene	11.079	95	99298	44.790	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.580%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.135	130	611	0.340	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037385.D
Acq On : 30 Aug 2023 13:39
Operator : JC/MD
Sample : 04150-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D1-20230822

Quant Time: Aug 31 02:22:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

