

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037389.D
 Acq On : 30 Aug 2023 15:17
 Operator : JC/MD
 Sample : 04229-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-07-6.5-082923

Quant Time: Aug 31 02:24:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243563	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	211744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	90522	42.308	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.620%
35) Dibromofluoromethane	5.385	113	74456	46.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.647	98	282328	48.251	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	11.079	95	100064	43.517	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.040%
Target Compounds						
16) Acetone	2.386	43	2385	2.849	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037389.D
Acq On : 30 Aug 2023 15:17
Operator : JC/MD
Sample : 04229-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-07-6.5-082923

Quant Time: Aug 31 02:24:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

