

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037392.D
 Acq On : 30 Aug 2023 16:30
 Operator : JC/MD
 Sample : 04229-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-02-9-082923

Quant Time: Aug 31 02:25:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	136986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	226291	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199135	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83679	42.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.740%
35) Dibromofluoromethane	5.385	113	68735	45.864	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.720%
50) Toluene-d8	8.647	98	264242	48.607	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.220%
62) 4-Bromofluorobenzene	11.079	95	94959	44.449	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.900%
Target Compounds						
					Qvalue	
24) 1,1-Dichloroethane	3.605	63	14446	5.143	ug/l	96
44) Trichloroethene	7.129	130	640	0.370	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037392.D
Acq On : 30 Aug 2023 16:30
Operator : JC/MD
Sample : 04229-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-02-9-082923

Quant Time: Aug 31 02:25:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

