

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037410.D
 Acq On : 31 Aug 2023 06:39
 Operator : JC/MD
 Sample : 04128-14 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP10-20230822

Quant Time: Aug 31 07:41:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145561	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	239491	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213638	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87467	41.680	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.360%
35) Dibromofluoromethane	5.385	113	71882	45.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.640%
50) Toluene-d8	8.647	98	277618	48.253	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.500%
62) 4-Bromofluorobenzene	11.079	95	103053	45.579	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.160%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.782	84	1573	0.857	ug/l	92
44) Trichloroethene	7.135	130	450	0.246	ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037410.D
Acq On : 31 Aug 2023 06:39
Operator : JC/MD
Sample : 04128-14 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP10-20230822

Quant Time: Aug 31 07:41:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

