

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037411.D
 Acq On : 31 Aug 2023 07:03
 Operator : JC/MD
 Sample : 04128-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE132D1-20230821

Quant Time: Aug 31 09:23:03 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	155345	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	257157	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	226806	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100131	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	93946	41.948	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.900%
35) Dibromofluoromethane	5.385	113	78207	45.921	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.840%
50) Toluene-d8	8.647	98	298149	48.262	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.520%
62) 4-Bromofluorobenzene	11.079	95	107581	44.313	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.620%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	1834	1.117	ug/l	97
27) cis-1,2-Dichloroethene	4.483	96	1551	0.724	ug/l	82
30) Chloroform	5.092	83	3215	0.899	ug/l	94
44) Trichloroethene	7.123	130	88986	45.223	ug/l	94
64) Tetrachloroethene	9.275	164	2861	1.729	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037411.D
Acq On : 31 Aug 2023 07:03
Operator : JC/MD
Sample : 04128-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE132D1-20230821

Quant Time: Aug 31 09:23:03 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

