

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037418.D
 Acq On : 31 Aug 2023 09:51
 Operator : JC/MD
 Sample : 04128-12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-MW01D3-20230822

Quant Time: Aug 31 18:23:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154168	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	255138	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110274	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95326	42.889	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.780%
35) Dibromofluoromethane	5.385	113	79145	46.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.680%
50) Toluene-d8	8.647	98	300736	49.066	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.140%
62) 4-Bromofluorobenzene	11.079	95	112273	46.611	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.220%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037418.D
Acq On : 31 Aug 2023 09:51
Operator : JC/MD
Sample : 04128-12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-MW01D3-20230822

Quant Time: Aug 31 18:23:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

