

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037420.D
 Acq On : 31 Aug 2023 10:40
 Operator : JC/MD
 Sample : 04128-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT205S1-20230822

Quant Time: Aug 31 18:23:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	250802	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	110254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94360	43.251	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.500%
35) Dibromofluoromethane	5.385	113	78216	47.090	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.180%
50) Toluene-d8	8.647	98	297980	49.457	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.079	95	112136	47.359	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.720%
Target Compounds						
					Qvalue	
44) Trichloroethene	7.123	130	9403	4.900	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037420.D
Acq On : 31 Aug 2023 10:40
Operator : JC/MD
Sample : 04128-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT205S1-20230822

Quant Time: Aug 31 18:23:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

