

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
 Data File : VX037423.D
 Acq On : 31 Aug 2023 11:52
 Operator : JC/MD
 Sample : 04128-18
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT180D1-20230821

Quant Time: Aug 31 18:24:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 18 09:30:38 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	151799	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	256193	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	233755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	95523	43.649	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	87.300%		
35) Dibromofluoromethane	5.391	113	78139	46.053	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	92.100%		
50) Toluene-d8	8.647	98	301437	48.978	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.960%		
62) 4-Bromofluorobenzene	11.079	95	116058	47.984	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	95.960%		
Target Compounds					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	6201	3.864	ug/l	99
12) 1,1-Dichloroethene	2.319	96	4322	2.766	ug/l	86
27) cis-1,2-Dichloroethene	4.495	96	3356	1.603	ug/l	79
30) Chloroform	5.092	83	4654	1.332	ug/l	93
44) Trichloroethene	7.123	130	484137	246.964	ug/l	98
55) 1,1,2-Trichloroethane	9.153	97	2328	1.252	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037423.D
Acq On : 31 Aug 2023 11:52
Operator : JC/MD
Sample : 04128-18
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT180D1-20230821

Quant Time: Aug 31 18:24:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 18 09:30:38 2023
Response via : Initial Calibration

