


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX083023\
Data File : VX037401.D
Acq On : 31 Aug 2023 02:14
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081723W.M
Title : SW846 8260
Last Update : Fri Aug 18 09:30:38 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	15.7	6936	PASS
75	95	30	60	49.2	21725	PASS
95	95	100	100	100.0	44139	PASS
96	95	5	9	6.8	2997	PASS
173	174	0.00	2	0.8	254	PASS
174	95	50	100	72.1	31837	PASS
175	174	5	9	7.6	2416	PASS
176	174	95	101	95.7	30459	PASS
177	176	5	9	6.4	1964	PASS