

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083118\
 Data File : VX004330.D
 Acq On : 31 Aug 2018 17:00
 Operator : JC/MD
 Sample : VX0831MBL01
 Misc : 5.00g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0831MBL01

Quant Time: Sep 01 05:07:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	265520	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	426028	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	397100	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	227021	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193972	56.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.10%	
35) Dibromofluoromethane	5.50	113	167065	52.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.42%	
50) Toluene-d8	8.71	98	607841	50.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.14	95	210031	47.05	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.10%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083118\
Data File : VX004330.D
Acq On : 31 Aug 2018 17:00
Operator : JC/MD
Sample : VX0831MBL01
Misc : 5.00µl/10mL/100µL/5.0mL/MSVOA_X/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0831MBL01

Quant Time: Sep 01 05:07:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

