

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083118\
 Data File : VX004334.D
 Acq On : 31 Aug 2018 18:47
 Operator : JC/MD
 Sample : J4275-19ME
 Misc : 4.64g/5mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CL-02-082918-AME

Quant Time: Sep 01 05:35:58 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 05:39:44 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	252348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	406512	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	400888	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	265385	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	181630	55.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.44%	
35) Dibromofluoromethane	5.50	113	153603	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
50) Toluene-d8	8.71	98	592879	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
62) 4-Bromofluorobenzene	11.14	95	238932	56.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.20%	
Target Compounds						
16) Acetone	2.48	43	6259	Below Cal		92
18) Methyl Acetate	2.79	43	14007	2.87 ug/l		99
20) Methylene Chloride	2.84	84	3878	1.22 ug/l		88
95) Naphthalene	13.83	128	29088	1.68 ug/l		99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083118\
Data File : VX004334.D
Acq On : 31 Aug 2018 18:47
Operator : JC/MD
Sample : J4275-19ME
Misc : 4.64g/5mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CL-02-082918-AME

Quant Time: Sep 01 05:35:58 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082218W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 05:39:44 2018
Response via : Initial Calibration

