

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018140.D
 Acq On : 31 Aug 2020 12:28
 Operator : JC/SP
 Sample : L3846-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB181-EB-20200828

Quant Time: Sep 01 05:59:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	615848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1009423	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	945717	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	434572	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	421278	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	5.46	113	341842	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
50) Toluene-d8	8.70	98	1250735	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
62) 4-Bromofluorobenzene	11.12	95	445154	44.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.38%	
Target Compounds						
3) Chloromethane	1.31	50	6258	0.792	ug/l	94
16) Acetone	2.42	43	32303	9.696	ug/l	99
17) Carbon Disulfide	2.55	76	5167	0.284	ug/l #	75
52) Toluene	8.76	92	45655	2.632	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018140.D
Acq On : 31 Aug 2020 12:28
Operator : JC/SP
Sample : L3846-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB181-EB-20200828

Quant Time: Sep 01 05:59:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

