

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX083120\
 Data File : VX018150.D
 Acq On : 31 Aug 2020 16:16
 Operator : JC/SP
 Sample : L3831-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 INTERCEPTOR-20200827

Quant Time: Sep 01 06:29:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 21 03:03:56 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	501771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	824906	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	763778	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	375282	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	353736	50.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.56%	
35) Dibromofluoromethane	5.47	113	281953	50.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.38%	
50) Toluene-d8	8.70	98	998904	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	11.12	95	371171	45.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.18%	
Target Compounds						
3) Chloromethane	1.31	50	2426	0.377	ug/l #	88
7) Trichlorofluoromethane	1.92	101	20166	2.307	ug/l	91
16) Acetone	2.42	43	12747	4.696	ug/l	96
27) cis-1,2-Dichloroethene	4.56	96	32333	5.165	ug/l	89
44) Trichloroethene	7.19	130	211966	33.554	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX083120\
Data File : VX018150.D
Acq On : 31 Aug 2020 16:16
Operator : JC/SP
Sample : L3831-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
INTERCEPTOR-20200827

Quant Time: Sep 01 06:29:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X082120W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 21 03:03:56 2020
Response via : Initial Calibration

